



# CERTIFICATE OF ACCREDITATION

**The ANSI National Accreditation Board**

Hereby attests that

**Fox Valley Metrology, Ltd.**  
3114 Medalist Drive  
Oshkosh, WI 54902  
(and satellite locations as shown on the scope)

Fulfills the requirements of

**ISO/IEC 17025:2017**

and national standards

**ANSI/NCSL Z540-1-1994 (R2002) and**  
**ANSI/NCSL Z540.3-2006 (R2013)**

In the fields of

## CALIBRATION AND DIMENSIONAL MEASUREMENT

This certificate is valid only when accompanied by a current scope of accreditation document.

The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

Jason Stine, Vice President

Expiry Date: 15 June 2025

Certificate Number: ACT-1272



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



## SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

**ANSI/NCSL Z540-1-1994 (R2002)**  
**ANSI/NCSL Z540.3-2006 (R2013)**

**Fox Valley Metrology, Ltd.**  
3114 Medalist Drive  
Oshkosh, WI 54902  
Brian Gliszinski 920-426-5894

Services performed at satellite locations as indicated in far-right column

308 Axminster Drive, Fenton, MO 63026  
30447 Stacy Ponds Drive, Stacy, MN 55079  
5245 27<sup>th</sup> Avenue, Rockford, IL 61109  
3012 Old Charlotte Hwy., Monroe, NC 28110  
1740 State Route 61, Crestline, OH 44827  
2205 North Willow Avenue, Unit B, Broken Arrow, OK 74012

## CALIBRATION AND DIMENSIONAL MEASUREMENT

Valid to: **June 15, 2025**

Certificate Number: **ACT-1272**

### CALIBRATION

#### Acoustics and Vibration

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Sound Level – Source <sup>1</sup> (100 Hz, 250 Hz, 500 Hz, 1 000 Hz, 2 000 Hz)	114 dB	0.6 dB	Gen Rad 1562-A Sound Level Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Accelerometers 1 g reference 1 g reference	10 Hz to 2 kHz (2 to 10) kHz	1.5 % of reading 1.4 % of reading	PCB 9150C Accelerometer Calibration Workstation  Oshkosh, WI

## Chemical Quantities

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH Meters <sup>1,6</sup>	(4.01, 7, 10) pH	0.02 pH	pH Buffer Solutions  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Conductivity Meters <sup>1,6</sup>	12.85 mS/cm 1 408 $\mu$ S/cm 10 $\mu$ S/cm	0.18 mS/cm 14 $\mu$ S/cm 0.18 $\mu$ S/cm	Conductivity Solutions  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Refractometers <sup>1,6</sup>	(4.99, 7.52, 10.03, 12.53, 14.98, 30.08) Brix	0.24 Brix	Refractive Index Solutions  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Gas Detectors <sup>1,6</sup> O <sub>2</sub>	0 % Concentration 21.9 % Concentration 100 % Concentration	2.4 % Concentration 2.4 % Concentration 2.6 % Concentration	Accredited Gas Mixtures  Oshkosh, WI

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source <sup>1,6</sup> (Fixed Value)	10 V	0.8 $\mu$ V/V	Fluke 732B Voltage Standard  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
DC Voltage – Source <sup>1</sup>	Up to 220 mV 220 mV to 2.2 V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1 100) V	12 $\mu$ V/V + 0.4 $\mu$ V 5.8 $\mu$ V/V + 0.7 $\mu$ V 4.2 $\mu$ V/V + 2.5 $\mu$ V 4.1 $\mu$ V/V + 4 $\mu$ V 5.8 $\mu$ V/V + 40 $\mu$ V 7.6 $\mu$ V/V + 0.4 mV	Fluke 5720A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
DC Voltage – Measure <sup>1</sup>	Up to 200 mV 200 mV to 2 V (2 to 20) V (20 to 200) V (200 to 1 050) V	5 $\mu$ V/V + 0.1 $\mu$ V 3.5 $\mu$ V/V + 0.4 $\mu$ V 3.5 $\mu$ V/V + 4 $\mu$ V 5.5 $\mu$ V/V + 40 $\mu$ V 5.5 $\mu$ V/V + 0.5 mV	Fluke 8508A 8.5 Digit Multimeter  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
DC High Voltage – Measure <sup>1</sup>	(1 to 10) kV (10 to 100) kV	60 V 600 V	Hipotronics KVM-100 High Voltage Meter  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Source <sup>1</sup>	Up to 220 µA 220 µA to 2.2 mA (2.2 to 22) mA (22 to 220) mA 220 mV to 2.2 A	0.12 mA/A + 6 nA 42 µA/A + 7 nA 41 µA/A + 40 nA 52 µA/A + 0.7 µA 93 µA/A + 12 µA	Fluke 5720A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
DC Current – Source <sup>1</sup>	(2.2 to 11) A (11 to 20.5) A	0.5 mA/A + 0.5 mA 1 mA/A + 0.75 mA	Fluke 5522A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
DC Current – Source <sup>1</sup>	Up to 100 A	0.008 % of reading + 4 mA	Fluke 52120A Transconductance Amplifier  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
DC Current – Source <sup>1</sup> Clamp-on Meters	Up to 2 500 A	0.6 % of reading	Fluke 52120A Transconductance Amplifier with 25-turn coil  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Measure <sup>1</sup>	Up to 200 µA 200 µA to 2 mA (2 to 20) mA (20 to 200) mA 200 mA to 2 A (2 to 20) A	12 µA/A + 0.4 nA 12 µA/A + 4 nA 14 µA/A + 40 nA 48 µA/A + 0.8 µA 0.19 mA/A + 16 µA 4 mA/A + 0.4 mA	Fluke 8508A 8.5 Digit Multimeter  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Resistance – Measure <sup>1</sup> (Normal Mode)	Up to 2 Ω (2 to 20) Ω (20 to 200) Ω 200 Ω to 2 kΩ (2 to 20) kΩ (20 to 200) kΩ 200 kΩ to 2 MΩ (2 to 20) MΩ (20 to 200) MΩ	17 µΩ/Ω + 4 µΩ 9.5 µΩ/Ω + 14 µΩ 8 µΩ/Ω + 50 µΩ 8 µΩ/Ω + 0.5 mΩ 8 µΩ/Ω + 5 mΩ 8 µΩ/Ω + 50 mΩ 9 µΩ/Ω + 1 Ω 20 µΩ/Ω + 0.1 kΩ 0.12 mΩ/Ω + 10 kΩ	Fluke 8508A 8.5 Digit Multimeter  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Resistance – Measure <sup>1</sup> High Voltage Mode	(2 to 20) MΩ (20 to 200) MΩ 200 mΩ to 2 GΩ (2 to 20) GΩ	17 µΩ/Ω + 10 Ω 65 µΩ/Ω + 1 kΩ 0.18 mΩ/Ω + 0.1 MΩ 15 mΩ/Ω + 10 MΩ	Fluke 8508A 8.5 Digit Multimeter  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Source <sup>1,6</sup> (Simulation-Fixed)	0 Ω 1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 kΩ 1.9 kΩ 10 kΩ 19 kΩ 100 kΩ 190 kΩ 1 MΩ 1.9 MΩ	0.11 mΩ 0.11 mΩ 0.21 mΩ 0.27 mΩ 0.51 mΩ 1.4 mΩ 2.6 mΩ 11 mΩ 21 mΩ 0.11 Ω 0.21 Ω 1.3 Ω 2.7 Ω 24 Ω 48 Ω	Fluke 5720A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Resistance – Source <sup>1,6</sup> (Simulation-Fixed)	10 MΩ 19 MΩ 100 MΩ	0.48 kΩ 1.1 kΩ 23 kΩ	Fluke 5720A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Resistance – Source <sup>1,6</sup> (Artifact-Fixed)	1 GΩ 10 GΩ 100 GΩ	1.9 MΩ 47 MΩ 0.95 GΩ	IET Labs HRRS Decade Box  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source <sup>1</sup>	Up to 2.2 mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (2.2 to 22) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (22 to 220) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (0.22 to 2.2) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.24 mV/V + 4 µV 0.09 mV/V + 4 µV 0.08 mV/V + 4 µV 0.2 mV/V + 4 µV 0.5 mV/V + 5 µV 1.05 mV/V + 10 µV 1.4 mV/V + 20 µV 2.7 mV/V + 20 µV  0.24 mV/V + 4 µV 0.09 mV/V + 4 µV 0.08 mV/V + 4 µV 0.2 mV/V + 4 µV 0.5 mV/V + 5 µV 1.05 mV/V + 10 µV 1.4 mV/V + 20 µV 2.7 mV/V + 20 µV  0.24 mV/V + 12 µV 0.09 mV/V + 7 µV 0.08 mV/V + 7 µV 0.2 mV/V + 7 µV 0.46 mV/V + 17 µV 0.9 mV/V + 20 µV 1.4 mV/V + 25 µV 2.7 mV/V + 45 µV  0.2 mV/V + 40 µV 0.075 mV/V + 15 µV 0.025 mV/V + 8 µV 0.055 mV/V + 10 µV 0.080 mV/V + 30 µV 0.23 mV/V + 80 µV 0.7 mV/V + 0.2 mV 1 mV/V + 0.3 mV	Fluke 5720A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source <sup>1</sup>	(2.2 to 22) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz  (22 to 220) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz  (220 to 1 100) V (15 to 50) Hz 50 Hz to 1 kHz	0.2 mV/V + 0.4 mV 0.075 mV/V + 0.15 mV 0.025 mV/V + 50 µV 0.055 mV/V + 0.1 mV 0.08 mV/V + 0.2 mV 0.25 mV/V + 0.6 mV 0.7 mV/V + 2 mV 1.1 mV/V + 3.2 mV  0.24 mV/V + 4 mV 0.09 mV/V + 1.5 mV 0.052 mV/V + 0.6 mV 0.08 mV/V + 1 mV 0.15 mV/V + 2.5 mV 0.9 mV/V + 16 mV 4.4 mV/V + 40 mV 8 mV/V + 80 mV  0.3 mV/V + 16 mV 0.07 mV/V + 3.5 mV	Fluke 5720A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
AC Voltage – Source <sup>1</sup>	(330 to 1 020) V 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.3 mV/V + 10 mV 0.25 mV/V + 10 mV 0.3 mV/V + 10 mV	Fluke 5522A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
AC Voltage Harmonics – Source <sup>1</sup>	(2 <sup>nd</sup> to 50 <sup>th</sup> ) (10 to 45) Hz (45 to 65) Hz (65 to 500) Hz 500 Hz to 5 kHz (5 to 10) kHz	32 mV to 33 V 33 mV to 1 000 V 33 mV to 1 000 V 330 mV to 1 000 V (3.3 to 1 000) V	Fluke 5522A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(1 to 10) kV (50 to 60) Hz (10 to 100) kV (50 to 60) Hz	0.12 kV 1.2 kV	Hipotronics KVM-100 High Voltage Meter  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
AC Voltage – Measure <sup>1</sup> Bandwidth < 1 MHz	Up to 200 mV (1 to 10) Hz (10 to 40) Hz (40 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz 200 mV to 2 V (1 to 10) Hz (10 to 40) Hz (40 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (2 to 20) V (1 to 10) Hz (10 to 40) Hz (40 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz	0.17 mV/V + 14 µV 0.14 mV/V + 4 µV 0.12 mV/V + 4 µV 0.11 mV/V + 2 µV 0.14 mV/V + 4 µV 0.34 mV/V + 8 µV 0.77 mV/V + 20 µV  0.15 mV/V + 0.12 mV 0.12 mV/V + 20 µV 90 µV/V + 20 µV 75 µV/V + 20 µV 0.11 mV/V + 20 µV 0.22 mV/V + 0.84 mV 0.57 mV/V + 0.2 mV 3 mV/V + 2 mV 10 mV/V + 2 mV  0.15 mV/V + 1.2 mV 0.12 mV/V + 0.2 mV 90 µV/V + 0.2 mV 75 µV/V + 0.2 mV 0.11 mV/V + 0.2 mV 0.22 mV/V + 8.4 mV 0.57 mV/V + 2 mV 3 mV/V + 20 mV 10 mV/V + 20 mV	Fluke 8508A 8.5 Digit Multimeter  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup> Bandwidth < 1 MHz	(20 to 200) V (1 to 10) Hz (10 to 40) Hz (40 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (200 to 1 050) V (1 to 10) Hz (10 to 40) Hz 40 Hz to 10 kHz (10 to 30) kHz (30 to 100) kHz	0.15 mV/V + 12 mV 0.12 mV/V + 2 mV 90 $\mu$ V/V + 2 mV 75 $\mu$ V/V + 2 mV 0.11 mV/V + 2 mV 0.22 mV/V + 84 mV 0.57 mV/V + 20 mV 3 mV/V + 0.2 V 10 mV/V + 0.2 V  0.15 mV/V + 70 mV 0.12 mV/V + 20 mV 0.12 mV/V + 20 mV 0.23 mV/V + 40 mV 0.58 mV/V + 0.2 V	Fluke 8508A 8.5 Digit Multimeter  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
AC Current – Source <sup>1</sup>	Up to 220 $\mu$ A (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz 220 $\mu$ A to 2.2 mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.30 mA/A + 16 nA 0.20 mA/A + 10 nA 0.16 mA/A + 8 nA 0.22 mA/A + 12 nA 1.3 mA/A + 65 nA  0.31 mA/A + 40 nA 0.22 mA/A + 35 nA 0.15 mA/A + 35 nA 0.24 mA/A + 0.11 $\mu$ A 1.3 mA/A + 0.65 $\mu$ A  0.32 mA/A + 0.4 $\mu$ A 0.23 mA/A + 0.35 $\mu$ A 0.15 mA/A + 0.35 $\mu$ A 0.24 mA/A + 0.55 $\mu$ A 1.3 mA/A + 5 $\mu$ A  0.3 mA/A + 4 $\mu$ A 0.2 mA/A + 3.5 $\mu$ A 0.15 mA/A + 2.5 $\mu$ A 0.24 mA/A + 3.5 $\mu$ A 1.3 mA/A + 10 $\mu$ A	Fluke 5720A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source <sup>1</sup>	220 mA to 2.2 A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (2 to 3) A (10 to 45) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.31 mA/A + 35 µA 0.53 mA/A + 80 µA 8.1 mA/A + 0.16 mA  2.1 mA/A + 0.1 mA 0.75 mA/A + 0.1 mA 6.9 mA/A + 1 mA 29 mA/A + 5 mA	Fluke 5720A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
AC Current – Source <sup>1</sup>	(3 to 11) A (45 to 100) Hz (0.1 to 1) kHz (1 to 5) kHz (11 to 20.5) A (45 to 100) Hz (0.1 to 1) kHz (1 to 5) kHz	60 µA/A + 2 mA 1 mA/A + 2 mA 30 mA/A + 2 mA  1.2 mA/A + 5 mA 1.5 mA/A + 5 mA 30 mA/A + 5 mA	Fluke 5522A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
AC Current – Source <sup>1</sup>	Up to 50 A (6 to 10) kHz Up to 100 A (3 to 6) kHz Up to 300 A (1 to 3) kHz Up to 1 000 A 300 Hz to 1 kHz Up to 2 500 A (10 to 300) Hz	0.8 % of reading 0.75 % of reading 0.7 % of reading 0.8 % of reading 0.6 % of reading	Fluke 52120A Transconductance Amplifier with 25-turn Coil  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
AC Current – Source <sup>1</sup>	Up to 120 A (10 to 65) Hz (65 to 300) Hz 300 Hz to 1 kHz	0.012 % of reading + 19.2 mA 0.023 % of reading + 27.6 mA 0.078 % of reading + 93.6 mA	Fluke 52120A Transconductance Amplifier  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current Harmonics – Source <sup>1</sup>	(2 <sup>nd</sup> to 50 <sup>th</sup> ) (10 to 45) Hz (45 to 65) Hz (65 to 500) Hz 500 Hz to 5 kHz (5 to 10) kHz	3.3 mA to 3 A 3.3 mA to 20.5 A 33 mA to 20.5 A 33 mA to 20.5 A (33 to 330) mA	1.1 mA/A + 4 µA 0.5 mA/A + 4 µA 1.2 mA/A + 0.1 mA 2.3 mA/A + 0.2 mA 4.6 mA/A + 0.4 mA
AC Current – Measure <sup>1</sup>	Up to 200 µA (1 to 10) Hz 10 Hz to 10 kHz (10 to 30) kHz (30 to 100) kHz 200 µA to 2 mA (1 to 10) Hz 10 Hz to 10 kHz (10 to 30) kHz (30 to 100) kHz (2 to 20) mA (1 to 10) Hz 10 Hz to 10 kHz (10 to 30) kHz (30 to 100) kHz (20 to 200) mA (1 to 10) Hz 10 Hz to 10 kHz (10 to 30) kHz 200 mA to 2 A 10 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (2 to 20) A 10 Hz to 2 kHz (2 to 10) kHz	0.31 mA/A + 20 nA 0.3 mA/A + 20 nA 0.71 mA/A + 20 nA 4 mA/A + 20 nA  0.31 mA/A + 0.2 µA 0.3 mA/A + 0.2 µA 0.71 mA/A + 0.2 µA 4 mA/A + 0.2 µA  0.31 mA/A + 2 µA 0.3 mA/A + 2 µA 0.71 mA/A + 2 µA 4 mA/A + 2 µA  0.31 mA/A + 20 µA 0.3 mA/A + 20 µA 0.63 mA/A + 20 µA  0.62 mA/A + 0.2 mA 0.73 mA/A + 0.2 mA 3 mA/A + 0.2 mA  0.82 mA/A + 2 mA 2.5 mA/A + 2 mA	Fluke 5522A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Measure <sup>1</sup> 42 Hz to 5 MHz	0.32 pF to 370 mF	1.1 mF/F	Hioki 3532-50 LCR Meter  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Capacitance – Source <sup>1</sup> (Simulation)	130 pF to 3.3 nF (3.3 to 11) nF (11 to 110) nF (110 to 330) nF 330 nF to 1.1 µF (1.1 to 3.3) µF (3.3 to 11) µF (11 to 33) µF (33 to 110) µF (110 to 330) µF 330 µF to 1.1 mF (1.1 to 3.3) mF (3.3 to 11) mF (11 to 33) mF (33 to 110) mF	0.5 % of reading + 10 pF 0.5 % of reading + 10 pF 0.5 % of reading + 0.1 nF 0.25 % of reading + 0.3 nF 0.25 % of reading + 1 nF 0.25 % of reading + 3 nF 0.25 % of reading + 10 nF 0.4 % of reading + 30 nF 0.45 % of reading + 0.1 µF 0.45 % of reading + 0.3 µF 0.45 % of reading + 1 µF 0.45 % of reading + 3 µF 0.45 % of reading + 10 µF 0.75 % of reading + 30 µF 1.1 % of reading + 0.1 mF	Fluke 5522A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Phase – Measure <sup>1</sup>	Up to 360° 10 Hz to 2 kHz (2 to 5) kHz (5 to 10) kHz (10 to 50) kHz (50 to 60) kHz (60 to 70) kHz (70 to 80) kHz (80 to 90) kHz (90 to 100) kHz (100 to 500) kHz 500 kHz to 1 MHz	0.026° 0.036° 0.048° 0.059° 0.07° 0.082° 0.093° 0.1° 0.12° 0.58° 1.2°	Clark Hess 6000A Phase Meter  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Power – Source <sup>1</sup>	10 mW to 330 W 330 W to 3 kW (3 to 20.5) kW	0.27 mW/W 0.26 mW/W 0.82 mW/W	Fluke 5522A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
AC Power – Source <sup>1</sup>	100 µW to 9 W (9 to 33) W (33 to 90) W (90 to 330) W (330 to 900) W 900 W to 2.2 kW	1.7 mW/W 1.2 mW/W 1.7 mW/W 1.2 mW/W 11 mW/W 4.6 mW/W	Fluke 5522A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Oscilloscopes <sup>1</sup>			
DC Voltage (50 Ω)	1 mV to 6.6 V	2.9 mV/V + 40 µV	Fluke 5522A SC1100 Multiproduct Calibrator
DC Voltage (1 MΩ)	1 mV to 130 V	0.55 mV/V + 40 µV	Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
AC Voltage (50 Ω)	1 mVp-p to 6.6 Vp-p	2.9 mV/V + 40 µV	
AC Voltage (1 MΩ)	1 mVp-p to 130 Vp-p	1.1 mV/V + 40 µV	
Leveled Sine Wave 50 kHz to 1.1 GHz	5 mVp-p to 5.5 Vp-p	51 mV/V + 0.1 mV	
Time Markers	1 ns to 5 s	6.4 µs/s	

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes <sup>1</sup>			
Wave Generator (50 Ω)	1.8 mVp-p to 2.5 Vp-p	35 mV/V + 0.1 mV	Fluke 5522A SC1100 Multiproduct Calibrator
Wave Generator (1 MΩ)	1.8 mVp-p to 55 Vp-p	35 mV/V + 0.1 mV	Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Pulse Generator – Width	(4 to 45) ns (45 to 500) ns	58 ms/s + 0.5 ns 58 ms/s + 4 ns	
Pulse Generator – Period	200 ns to 20 ms	58 ms/s + 0.2 μs	
Input Impedance Measure	(50 to 60) Ω 500 kΩ to 1 MΩ	1.2 mΩ/Ω 1.2 mΩ/Ω	
Electrical Simulation of RTD Indicating Devices <sup>1</sup>	Pt 385, 100 Ω	0.05 °C	
	(-200 to -80) °C	0.05 °C	
	(-80 to 0) °C	0.05 °C	
	(0 to 100) °C	0.07 °C	
	(100 to 300) °C	0.09 °C	
	(300 to 400) °C	0.1 °C	
	(500 to 630) °C	0.12 °C	
	(630 to 800) °C	0.23 °C	
	Pt 3926, 100 Ω		Fluke 5522A Multiproduct Calibrator
	(-200 to -80) °C	0.05 °C	Oshkosh, WI
	(-80 to 0) °C	0.05 °C	Fenton, MO
	(0 to 100) °C	0.07 °C	Stacy, MN
	(100 to 300) °C	0.09 °C	Rockford, IL
	(300 to 400) °C	0.1 °C	Monroe, NC
	(500 to 630) °C	0.12 °C	Crestline, OH
	Pt 3916 (JIS), 100 Ω		Tahlequah, OK
	(-200 to -190) °C	0.25 °C	
	(-190 to -80) °C	0.04 °C	
	(-80 to 0) °C	0.05 °C	
	(100 to 260) °C	0.06 °C	
	(260 to 300) °C	0.07 °C	
	(300 to 400) °C	0.09 °C	
	(400 to 600) °C	0.1 °C	
	(600 to 630) °C	0.23 °C	

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of RTD Indicating Devices <sup>1</sup>	Pt 385, 200 Ω (-200 to -80) °C	0.04 °C	
	(-80 to 0) °C	0.04 °C	
	(0 to 100) °C	0.04 °C	
	(100 to 260) °C	0.05 °C	
	(260 to 300) °C	0.12 °C	
	(300 to 400) °C	0.13 °C	
	(400 to 600) °C	0.14 °C	
	(600 to 630) °C	0.16 °C	
	Pt 385, 500 Ω (-200 to -80) °C	0.04 °C	
	(-80 to 0) °C	0.05 °C	
	(0 to 100) °C	0.05 °C	
	(100 to 260) °C	0.06 °C	
	(260 to 300) °C	0.08 °C	
	(300 to 400) °C	0.08 °C	
	(400 to 600) °C	0.09 °C	
	(600 to 630) °C	0.11 °C	
	Pt 385, 1 000 Ω (-200 to -80) °C	0.03 °C	
	(-80 to 0) °C	0.03 °C	
	(0 to 100) °C	0.04 °C	
	(100 to 260) °C	0.05 °C	
	(260 to 300) °C	0.06 °C	
	(300 to 400) °C	0.07 °C	
	(400 to 600) °C	0.07 °C	
	(600 to 630) °C	0.23 °C	
	PtNi 385, 120 Ω, Ni 120 (-80 to 0) °C	0.08 °C	
	(0 to 100) °C	0.08 °C	
	(100 to 260) °C	0.14 °C	
	Cu 427, 10 Ω (-100 to 260) °C	0.03 °C	

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure <sup>1</sup>	Type K (-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 1 000) °C (1 000 to 1 372) °C  Type J (-210 to -100) °C (-100 to -30) °C (-30 to 150) °C (150 to 760) °C (760 to 1 200) °C  Type E (-250 to -100) °C (-100 to -35) °C (-25 to 350) °C (350 to 650) °C (650 to 1 000) °C  Type T (-250 to -150) °C (-150 to 0) °C (0 to 120) °C (120 to 400) °C  Type S (0 to 250) °C (250 to 1 000) °C (1 000 to 1 400) °C (1 400 to 1 767) °C  Type B (600 to 800) °C (-100 to -25) °C (-25 to 120) °C (120 to 1 000) °C  Type C (0 to 150) °C (150 to 650) °C (650 to 1 000) °C (1 000 to 1 800) °C (1 800 to 2 316) °C	0.33 °C 0.18 °C 0.16 °C 0.26 °C 0.4 °C  0.27 °C 0.16 °C 0.14 °C 0.17 °C 0.23 °C  0.5 °C 0.16 °C 0.14 °C 0.16 °C 0.21 °C  0.63 °C 0.24 °C 0.16 °C 0.14 °C  0.47 °C 0.36 °C 0.37 °C 0.46 °C  0.44 °C 0.34 °C 0.3 °C 0.33 °C  0.3 °C 0.26 °C 0.31 °C 0.5 °C 0.84 °C	Fluke 5522A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure <sup>1</sup>	Type L (-200 to -100) °C (-100 to 800) °C (800 to 900) °C  Type N (-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 410) °C (410 to 1 300) °C  Type R (0 to 250) °C (250 to 400) °C (400 to 1 000) °C (1 000 to 1 767) °C  Type U (-200 to 0) °C (0 to 600) °C	0.37 °C 0.26 °C 0.17 °C  0.4 °C 0.22 °C 0.19 °C 0.18 °C 0.27 °C  0.57 °C 0.35 °C 0.33 °C 0.4 °C  0.56 °C 0.27 °C	Fluke 5522A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Inductance – Source <sup>1</sup> (Artifact-Variable)	(1 to 10) mH (10 to 100) mH 100 mH to 1 H (1 to 10) H	22 mH/H 11 mH/H 6 mH/H 3 mH/H	General Radio 1490-D Decade Inductor  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Ionizers <sup>1</sup> Decay Time Float Voltage	(0.1 to 999.9) s (-1 100 to 1 100) V	0.2 s 3.1 V	Trek 156A Charged Plate Monitor  Oshkosh, WI
ESD Simulators Rise Time Burst/Surge Peak Current 30 ns Current 60 ns Current RC Time Constant	700 ps to 1 ns (7.5 to 30) A (4 to 16) A (2 to 8) A 600 ns 300 ns	0.14 ns 50 mA/A 0.1 A/A 0.12 A/A 20 ns 15 ns	Tektronix TDS684B Oscilloscope with EM Test CTR2 ESD Target IEC 61000-4-2, SAE J1113-13  Oshkosh, WI

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Transient Generators Rise Time Open Circuit Closed Circuit	75 ns to 10 $\mu$ s 75 ns to 10 $\mu$ s	1.5 % of reading + 0.12 $\mu$ s 0.69 % of reading + 79 ns	Oscilloscope, High Voltage Differential Probe, Current Probe per IEC 61000-4-5, IEC 61000-4-12, IEC 61000-4-18.
Duration/Pulse Width Open Circuit Closed Circuit	(50 to 700) $\mu$ s (20 to 320) $\mu$ s	0.058 % of reading + 0.28 $\mu$ s 0.12 % of reading + 34 ns	
Peak Voltage	(0.5 to 6) kV	1.6 % of reading + 9.5 V	
Peak Current	12.5 A to 3 kA	2.2 % of reading + 0.2 A	Oshkosh, WI
Frequency	5 kHz to 1 MHz	0.12 % of reading + 1.6 Hz	
Defibrillators	Up to 360 J	0.41 % of reading + 0.77 J	Oscilloscope, Tektronix P6015 High Voltage Probe, Digital Multimeter, Power Resistor  Oshkosh, WI
Current Injection Probes	9 kHz to 400 MHz	0.22 dB	VNA, VNA Calibration Kit per IEC 61000-4-6.  Oshkosh, WI

## Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1,4</sup> Absolute Level	(20 to 30) dBm 100 kHz to 3 GHz (3 to 18) GHz (18 to 26.5) GHz (-20 to 20) dBm 100 kHz to 3 GHz (3 to 18) GHz (18 to 26.5) GHz	0.37 dB 0.39 dB 0.4 dB 0.15 dB 0.18 dB 0.21 dB	Agilent N5531S Measuring Receiver with N5532A Sensor Modules  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
RF Power – Measure <sup>1</sup> Relative Level	(3.05 to 6.6) GHz (-90 to +30) dBm (-113 to -90) dBm (6.6 to 13.2) GHz (-81 to +30) dBm (-104 to -81) dBm (13.2 to 19.2) GHz (-70 to +30) dBm (-93 to -70) dBm	0.026 dB + 0.005 dB/10 dB 0.067 dB + 0.12 dB/10 dB  0.026 dB + 0.005 dB/10 dB 0.062 dB + 0.12 dB/10 dB  0.026 dB + 0.005 dB/10 dB 0.056 dB + 0.12 dB/10 dB	Agilent N5531S Measuring Receiver with N5532A Sensor Modules  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
RF Power – Measure <sup>1</sup> Relative Level	(19.2 to 26.5) GHz (-62 to +30) dBm (-85 to -62) dBm	0.026 dB + 0.005 dB/10 dB 0.051 dB + 0.12 dB/10 dB	Agilent N5531S Measuring Receiver with N5532A Sensor Modules  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Amplitude Modulation – Source <sup>1,4</sup> 250 kHz to 40 GHz	Rate: DC to 100 kHz Depths: (0 to 100) %	7.1 % of reading	Agilent E8257D Signal Generator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Amplitude Modulation – Measure <sup>1,4</sup> 100 kHz to 10 MHz  10 MHz to 3 GHz  10 MHz to 3 GHz  (3 to 26.5) GHz  (3 to 26.5) GHz	Rate: 50 Hz to 10 kHz Depths: (5 to 99) %  Rate: 50 Hz to 100 kHz Depths: (20 to 99) %  Rate: 50 Hz to 100 kHz Depths: (5 to 20) %  Rate: 50 Hz to 100 kHz Depths: (20 to 99) %  Rate: 50 Hz to 100 kHz Depths: (5 to 20) %	2.2 % of reading  1.2 % of reading  4.2 % of reading  3.5 % of reading  6 % of reading	Agilent N5531S Measuring Receiver with N5532A Sensor Modules  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Phase Modulation – Source <sup>1,4</sup> Rate: DC to 100 kHz	250 kHz to 40 GHz	5.9 % of reading + 0.1 rad	Agilent E8257D Signal Generator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Tuned RF Level – Measure <sup>1,4</sup> Absolute Level	500 kHz to 3.05 GHz (16 to 30) dBm (-106 to 16) dBm (-129 to -106) dBm (3.05 to 6.6) GHz (20 to 30) dBm (-90 to 20) dBm (-114 to -90) dBm (6.6 to 13.2) GHz (20 to 30) dBm (-81 to 20) dBm (-104 to -81) dBm	0.37 dB + 0.005 dB/10 dB 0.15 dB + 0.005 dB/10 dB 0.15 dB + 0.12 dB/10 dB  0.39 dB + 0.005 dB/10 dB 0.18 dB + 0.005 dB/10 dB 0.23 dB + 0.12 dB/10 dB  0.39 dB + 0.005 dB/10 dB 0.18 dB + 0.005 dB/10 dB 0.23 dB + 0.12 dB/10 dB	Agilent N5531S Measuring Receiver with N5532A Sensor Modules  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Tuned RF Level – Measure <sup>1,4</sup> Absolute Level	(13.2 to 19.2) GHz (+20 to +30) dBm (-70 to +20) dBm (-93 to -70) dBm  (19.2 to 26.5) GHz (+20 to +30) dBm (-62 to +20) dBm (-85 to -62) dBm  500 kHz to 3.05 GHz (-90 to +30) dBm (-106 to -90) dBm (-129 to -106) dBm	0.4 dB + 0.005 dB/10 dB 0.21 dB + 0.005 dB/10 dB 0.25 dB + 0.12 dB/10 dB  0.4 dB + 0.005 dB/10 dB 0.21 dB + 0.005 dB/10 dB 0.24 dB + 0.12 dB/10 dB  0.026 dB + 0.005 dB/10 dB 0.067 dB + 0.12 dB/10 dB 0.076 dB + 0.12 dB/10 dB	Agilent N5531S Measuring Receiver with N5532A Sensor Modules  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
RF Power – Source <sup>1</sup>	> -10 dBm  250 kHz to 2 GHz (2 to 20) GHz (20 to 40) GHz (-10 to -70) dBm  250 kHz to 2 GHz (2 to 20) GHz (20 to 40) GHz (-70 to -90) dBm  250 kHz to 2 GHz (2 to 20) GHz (20 to 40) GHz	0.72 dB 0.96 dB 1.1 dB  0.89 dB 1.1 dB 1.2 dB  0.95 dB 1.2 dB 1.21 dB	Agilent E8257D Signal Generator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Sensors – Calibration Factor <sup>1,4</sup>	(-20 to +14) dBm 100 kHz to 10 MHz 10 MHz to 10 GHz (10 to 18) GHz	1.5 % of reading 1.5 % of reading 1.7 % of reading	Tegam 1827 Power Sensor Calibrator, Agilent 3458A 8.5 Digit Multimeter, Agilent E8257D Signal Generator, Agilent E4419B Power Meter, Agilent 3325B Function Generator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Frequency Modulation – Measure <sup>1,4</sup>	250 kHz to 10 MHz  10 MHz to 3 GHz  (3 to 26.5) GHz	Rate: 20 Hz to 10 kHz Dev.: $\leq 40$ kHz peak  Rate: 20 Hz to 200 kHz Dev.: $\leq 400$ kHz peak  Rate: 20 Hz to 200 kHz Dev.: $\leq 400$ kHz peak	3.1 % of reading 3.1 % of reading 7.7 % of reading
Frequency Modulation – Source <sup>1,4</sup>	250 kHz to 40 GHz	1 dB Rate: DC to 100 kHz 3 dB Rate: DC to 10 MHz	Agilent E8257D Signal Generator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pulse Generation – Measure <sup>1,4</sup> DC to 225 MHz Pulse Width Rise/Fall Time	5 ns to 1 000 000 s 5 ns to 1 000 000 s	1.1 ns 1.1 ns	Agilent 53132A Counter  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Pulse Generation – Source <sup>1,4</sup> Repetition Frequency: 24 mHz to 14.28 MHz Period: 70 ns to 42 s	10 ns to 42 s	1.7 ns	Agilent E8257D Signal Generator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Line Impedance Stabilization Network <sup>1,4</sup> Insertion Loss	(-20 to 0) dB 9 kHz to 400 MHz	0.25 dB	Vector Network Analyzer, Attenuators, VNA Cal Kit, ANSI C63.4, CISPR 25, CISPR 16-1-2, DO-160G, MIL-STD 461G
Impedance – Magnitude	100 mΩ to 1 kΩ 9 kHz to 400 MHz	5.6 % of reading	Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Impedance – Phase	(-180 to 180)° 9 kHz to 400 MHz	5.3°	
Isolation (De-coupling Factor)	(-90 to 0) dB 9 kHz to 400 MHz	0.37 dB	

### Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Coupling/De-coupling Network <sup>1,4</sup>			Vector Network Analyzer, Attenuators, VNA Cal Kit, IEC 61000-4-6, CISPR 16-1-2
Insertion Loss	(-20 to 0) dB 9 kHz to 230 MHz	0.25 dB	Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Impedance – Magnitude	100 mΩ to 1 kΩ 9 kHz to 230 MHz	5.6 % of reading	

### Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Gage Blocks <sup>2</sup>	Up to 30 in	$(3.8 + 0.93L) \mu\text{in}$	Mahr 828 Measuring Machine, Per ASME B89.1.9  Oshkosh, WI
Gage Blocks <sup>2</sup>	Up to 10 in	$(3.9 + 1.3L) \mu\text{in}$	P & W Universal Labmaster, Per ASME B89.1.9  Oshkosh, WI Rockford, IL
Gage Blocks <sup>2</sup>	Up to 20 in	$(8.5 + 1L) \mu\text{in}$	ULM 600 Measuring Machine, Per ASME B89.1.9  Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH

## Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Length Standards <sup>2</sup>	Up to 24 in	(12 + 1L) $\mu$ in	ULM 600 Measuring Machine, Per ASME B89.1.1  Fenton, MO Stacy, MN Rockford, IL Monroe, NC
Length Standards <sup>2</sup>	Up to 39 in	(12 + 1L) $\mu$ in	Mahr 828 Measuring Machine, Per ASME B89.1.1  Oshkosh, WI
Length Standards <sup>2</sup>	(39 to 70) in	(390 + 2.6L) $\mu$ in	CMM  Oshkosh, WI Stacy, MN
Length Standards <sup>2</sup>	Up to 4 in	(53 + 0.4L) $\mu$ in	Plug Gage Comparator  Crestline, OH Tahlequah, OK
Cylindrical Rings <sup>2</sup>	(0.02 to 18) in	(8 + 1.8D) $\mu$ in	Mahr 828 Measuring Machine, ASME B89.1.6  Oshkosh, WI
Cylindrical Rings <sup>1,2</sup>	(0.25 to 8) in	(12 + 3D) $\mu$ in	Fowler Lab Concept Measuring Machine, ASME B89.1.6  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Cylindrical Plugs <sup>2</sup>	Up to 30 in	(2.7 + 6D) $\mu$ in	Mahr 828 Measuring Machine  Oshkosh, WI

## Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Cylindrical Plugs <sup>1,2</sup>	Up to 4 in	(53 + 0.4D) $\mu$ in	Plug Gage Comparator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Thread Rings <sup>2</sup> Pitch Diameter Minor Diameter	Up to Setting plug size Up to 16 in Up to 9 in	(240 + 0.3D) $\mu$ in  120 $\mu$ in	Setting Plug Gages, Measuring Machine, ID Bore Gages, ASME B1.2  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Non-Standard Thread Rings <sup>2</sup> Pitch Diameter Minor Diameter	Up to 14 in Up to 9 in	(120 + 2.5D) $\mu$ in  120 $\mu$ in	Mahr ULM 600 Measuring Machine, ID Bore Gages, ASME B1.2  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC

## Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
NPT Rings (Standoff and Basic Length)	(0.062 5 to 8) in	250 $\mu$ in	NPT Plugs, P&W LabMaster, P&W Laser Ruler ASME B1.20.5 ASME B1.20.1  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH
NPT Plugs (Standoff and Basic Length)	(0.062 5 to 6) in	490 $\mu$ in	NPT Rings, P&W LabMaster, P&W Laser Ruler ASME B1.20.5 ASME B1.20.1  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH
Thread Plugs <sup>2</sup> Pitch Diameter Major Diameter	Up to 24 in	$(73 + 0.9D) \mu$ in	P&W Supermicrometer, Thread Measuring Wires, ASME B1.2  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC
Thread Plugs <sup>1,2</sup> Pitch Diameter Major Diameter	Up to 4 in	$(73 + 3.2D) \mu$ in	Plug Gage Comparator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Thread Wires <sup>2</sup>	Up to 0.5 in	(11 + 1.5D) $\mu$ in	Mahr ULM 600 Measuring Machine, ASME B89.1.17  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC
Calipers <sup>1,2</sup>	Up to 80 in	(380 + 15L) $\mu$ in	Gage Blocks  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Indicators <sup>1,2</sup>	Up to 4 in	(36 + 10L) $\mu$ in	Indicator Checker  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Test Indicators <sup>1</sup>	Up to 0.06 in	39 $\mu$ in	Indicator Checker  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
OD Micrometers <sup>1,2</sup>	Up to 60 in	(72 + 12L) $\mu$ in	Gage Blocks Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
ID Micrometer <sup>1,2</sup>	(1.5 to 40) in	(370 + 7L) $\mu$ in	Gage Blocks Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Height Gages <sup>1,2</sup>	Up to 40 in	(96 + 14L) $\mu$ in	Gage Blocks Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Bore Gages <sup>1</sup>	(0.25 to 12) in	45 $\mu$ in	Cylindrical Rings Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Crimpers <sup>1</sup>			
Die Check	(0.011 to 0.5) in	230 $\mu$ in	Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Crimp Height	(0.01 to 0.5) in	0.001 2 in	
Profilometers <sup>1</sup>	(2 to 300) $\mu$ m Ra	2.2 $\mu$ m	Roughness Specimen  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Surface Plates <sup>1,2</sup>			
Repeat Reading	(4 to 34) inDL	35 $\mu$ m	In accordance with ASME B89.3.7 using Repeat-O-Meter
Overall Flatness	(34 to 175) inDL	(92 + 0.14DL) $\mu$ m	Electronic Levels  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Pi Tapes Diameter	Up to 204 in	0.000 14 % of reading + 260 $\mu$ m	Renishaw Laser Measuring System w/Microscope  Oshkosh, WI
Profilometer Reference Specimens	(0.01 to 600) $\mu$ m Ra	2.4 $\mu$ m	Profilometer  Oshkosh, WI

## Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
CMM Calibration <sup>1,2</sup> Volumetric Linearity	(5 to 40) in  (1 to 60) in > 60 in	(12 + 14L) $\mu$ in  (7 + 14L) $\mu$ in (20 + 0.4L) $\mu$ in	Ball Bars, Step Gage, Renishaw Laser System, B89.4.10  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Linear Measurements	Up to 1 560 in	(38 + 0.5L) $\mu$ in	Renishaw Laser System  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Optical Comparators <sup>1,2</sup> Linearity of Table Travel Magnification	Up to 30 in  10x, 20x, 31.25x, 50x, 62.5x, 100x, 200x	(97 + 12L) $\mu$ in  460 $\mu$ in	Glass Scale, Precision Balls, Calibration Sphere  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Roundness Testers <sup>1</sup> Axial Error Radial Error	(-1 000 to 1 000) $\mu$ m  (-1 000 to 1 000) $\mu$ m	0.15 $\mu$ m  0.15 $\mu$ m	Test Sphere  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
ULMs <sup>1</sup> (Length)	Up to 100 mm	0.19 $\mu\text{m}$	Gage Blocks  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Thickness Gages <sup>1</sup>	Up to 0.06 in Up to 6 in	380 $\mu\text{in}$ 380 $\mu\text{in}$	Film Thickness Standards Gage Blocks  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Brinell Scopes <sup>1</sup>	(1 to 6) mm	11 $\mu\text{m}$	Stage Micrometer  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Analytical Balances <sup>1,5</sup>	Up to 20 g Up to 100 g Up to 10 kg	0.038 % of reading + 1.6 µg 0.077 % of reading + 1.6 µg 0.009 4 % of reading + 69 µg	ASTM E617 Class 1 Weights and NIST Handbook 44 utilized in the calibration of the weighing system.  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Bench and Floor Scales <sup>1,5</sup>	Up to 10 lb (>10 to 5 000) lb	1.3 mg 0.001 3 % of reading	NIST 105 Class F Weights and NIST Handbook 44 utilized in the calibration of the weighing system.  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Mass – Avoirdupois	50 lb 20 lb 5 lb 2 lb 1 lb 0.5 lb 1 oz	0.29 g 0.12 g 5.8 mg 5.8 mg 5.8 mg 5.8 mg 9.5 µg	Class 1 Weights and Analytical Balance, Modified Substitution  Oshkosh, WI Fenton, MO (Class F Only) Stacy, MN (Class F Only)

## Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Mass – SI	25 000 g 20 000 g 5 000 g 3 000 g 2 000 g 1 000 g 500 g 300 g 200 g 100 g 50 g 30 g 20 g 10 g 5 g 3 g 2 g 1 g	0.29 g 0.29 g 5.9 mg 5.8 mg 5.8 mg 5.8 mg 5.8 mg 5.8 mg 0.11 mg 97 µg 95 µg 94 µg 11 µg 9.3 µg 3.8 µg 2.8 µg 2.4 µg 2.4 µg	Class 1 Weights and Analytical Balance, Modified Substitution  Oshkosh, WI Fenton, MO (Class F Only) Stacy, MN (Class F Only)
Mass – SI	500 mg 200 mg 100 mg 50 mg 20 mg 10 mg 5 mg 3 mg 2 mg 1 mg	2 µg 1.6 µg 1.6 µg 1.6 µg 1.7 µg 1.7 µg 1.7 µg 1.8 µg 1.7 µg 1.6 µg	Class 1 Weights and Analytical Balance, Modified Substitution  Oshkosh, WI Fenton, MO (Class F Only) Stacy, MN (Class F Only)
Gauge Pressure <sup>1</sup>	Up to 854 inH <sub>2</sub> O	0.03 % of reading + 0.000 044 inH <sub>2</sub> O	Ametek PK2 Deadweight Tester  Oshkosh, WI
Absolute Pressure	(0.2 to 25) psia (25 to 500) psia	0.001 2 % of reading 0.002 7 % of reading + 0.000 4 psi	Ruska 2465 Deadweight Tester  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

### Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Gauge Pressure	(500 to 3 000) psig	0.003 % of reading + 0.000 9 psi	Ruska 2470 Deadweight Tester  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Gauge Pressure	Up to 600 psig (600 to 40 000) psig	0.005 3 % of reading + 0.002 1 psi 0.008 % of reading	Budenberg BGH2600 Deadweight Tester  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Gauge Pressure	(40 000 to 60 000) psig	36 psi	Additel Hydraulic Pump, Digital Pressure Test Gauge  Oshkosh, WI
Durometers			Full Direct Verification per ASTM D2240 using Durometer Calibrator, Triple Beam Balance
Spring Force	Up to 100 duro	0.1 duro	
Indenter Dimensions			
Indenter Angle	(20 to 40) <sup>o</sup>	0.11 <sup>o</sup>	
Indenter Radius	Up to 0.1 in	160 $\mu$ in	Video Measuring Machine
Indenter Length	Up to 0.198 in	22 $\mu$ in	Video Measuring Machine  Gage Blocks  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC

## Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Brinell Hardness Testers <sup>1</sup>  Verification of Test Force	(500, 750, 1 500, 2 000, 3 000) kgf	0.072 % of reading + 4.2 kgf	Partial Direct Verification per ASTM E10 using Morehouse Proving Ring  Video Measuring Machine  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Brinell Hardness Testers <sup>1</sup>	(50 to 650) HBW	1.2 % of reading + 3.2 HBW	Indirect Verification per ASTM E10 using Brinell Test Blocks and Brinell Scope  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Knoop and Vickers Hardness Testers <sup>1</sup>	HK0.05 (250 to 650) HK > 650 HK HK0.1 (250 to 650) HK > 650 HK HK0.3 (250 to 650) HK > 650 HK HK0.5 (250 to 650) HK > 650 HK HK1.0 (250 to 650) HK > 650 HK	11 HK 27 HK  11 HK 25 HK  11 HK 18 HK  14 HK 17 HK  11 HK 16 HK	Indirect Verification per ASTM E384 using Knoop and Vickers Test Blocks  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

### Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Knoop and Vickers Hardness Testers <sup>1</sup>	HV0.05 (250 to 650) HV > 650 HV HV0.1 (250 to 650) HV > 650 HV HV0.3 (250 to 650) HV > 650 HV HV0.5 (250 to 650) HV > 650 HV HV1.0 (250 to 650) HV > 650 HV	10 HK 39 HK 9 HK 30 HK 10 HK 18 HK 7 HK 17 HK 7 HK 14 HK	Indirect Verification per ASTM E384 using Knoop and Vickers Test Blocks  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Leeb Hardness Tester <sup>1</sup>	(300 to 900) LD	12 LD	Indirect Verification per ASTM A596 using Leeb Test Blocks  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Rockwell Hardness Testers <sup>1</sup>	HRBW Low HRBW Med HRBW High  HRA Low HRA Med HRA High  HRC Low HRC Med HRC High	0.77 HRBW 0.79 HRBW 0.71 HRBW  0.56 HRA 0.49 HRA 0.47 HRA  0.52 HRC 0.53 HRC 0.54 HRC	Indirect Verification per ASTM E18 using Rockwell Test Blocks  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Rockwell Hardness Testers <sup>1</sup>	HREW Low HREW Med HREW High  HRFW Low HRFW Med HRFW High  HRHW Low HRHW Med HRHW High  HR15N Low HR15N Med HR15N High  HR30N Low HR30N Med HR30N High  HR45N Low HR45N Med HR45N High HR15TW Low HR15TW Med HR15TW High  HR30TW Low HR30TW Med HR30T High  HR45TW Low HR45TW Med HR45TW High	0.72 HREW 0.7 HREW 0.64 HREW  0.71 HRFW 0.59 HRFW 0.57 HRFW  0.72 HRHW 0.58 HRHW 0.6 HRHW  0.72 HR15N 0.7 HR15N 0.55 HR15N  0.67 HR30N 0.66 HR30N 0.64 HR30N  0.58 HR45N 0.65 HR45N 0.62 HR45N 0.72 HR15TW 0.74 HR15TW 0.52 HR15TW  0.64 HR30TW 0.65 HR30TW 0.54 HR30TW  0.67 HR45TW 0.69 HR45TW 0.69 HR45TW	Indirect Verification per ASTM E18 using Rockwell Test Blocks  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Force Devices <sup>1</sup>	Up to 100 lb	0.05 % of reading	Dead Weight Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Force Devices <sup>1</sup>	Up to 500 lb (500 to 2 000) lb (2 000 to 12 500) lb (12 500 to 50 000) lb (50 000 to 500 000) lb	0.05 % of reading 0.07 % of reading 0.1 % of reading 0.1 % of reading 0.1 % of reading	Load Cells Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Force Devices	Up to 200 lb	0.05 % reading	Dead Weight Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Force Devices	Up to 1 000 lb (1 000 to 10 000) lb (10 000 to 100 000) lb	0.03 % reading 0.03 % reading 0.04 % reading	Load Cells Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Wedge Tester <sup>1</sup>	Up to 40 000 N	32 N	Load Cell  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Torque Transducers <sup>1</sup>	0.000 04 in oz to 250 lbf·ft (250 to 2 000) lbf·ft	0.05 % of reading 0.06 % of reading	Torque Arms, Dead Weight  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC
Torque Tools <sup>1</sup>	(10 to 200) ozf·in 4 lbf·in to 2 000 lbf·ft	0.5 % of reading 0.3 % of reading	CDI Torque System  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Viscosity Rotational Viscometers	500 cP 5 000 cP	0.02 cP/cP	Viscosity Solutions, Temperature Bath  Oshkosh, WI Stacy, MN
Viscosity Cups	17.82 cP 65.28 cP 231 cP	0.03 cP/cP	Viscosity Solutions, Temperature Bath, Stopwatch Per ASTM D4212  Oshkosh, WI Stacy, MN

## Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pipettes <sup>1</sup>	Up to 20 µL (20 to 200) µL (200 to 1 000) µL (1 000 to 5 000) µL (5 000 to 10 000) µL	0.064 µL 0.6 µL 1.8 µL 8.9 µL 18 µL	Analytical Balance Per ISO 8655-6  Oshkosh, WI
Liquid Volume Measuring Devices	Up to 200 mL (200 to 6 000) mL (6 000 to 34 000) mL	0.003 mL 0.46 mL 2.8 mL	Balances  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Foundry Sand Test Equipment / Measurement <sup>1</sup>			
Ultrasonic Cleaner/Scrubber	18 °F 30 m	1.7 °F 1.2 s	Temperature Calibrator Stopwatch
Wet Tensile Tester	0.449 N/cm <sup>2</sup> (300 to 320) °F	0.003 1 N/cm <sup>2</sup> 2 °F	Dead Weight Temperature Calibrator
Sand Squeezer	Up to 200 psi	3.8 psi	Proving Ring
Tensile Testers	Up to 10 000 lb	7.2 lb	Load Cell  Oshkosh, WI

## Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Foundry Sand Test Equipment / Measurement <sup>1</sup>			
AFS Clay Tester	Up to 10 min	1.2 s	Stopwatch
Friability Tester	60 s	1.2 s	Stopwatch
Sand Rammer	Up to 2 in	0.01 in	Impact Rings
Moisture Teller	(0 to 300 °F)	1.9 °F	Temperature Calibrator
Permmeter	Up to 500 perm	0.43 perm	Perm Standards
Sand Strength Tester	Up to 500 psi Up to 1 000 lb	1.1 psi 4.2 lb	Proving Ring
Core Scratch Tester	Up to 0.1 in	0.006 in	Flatness Block
Green Sand Hardness Tester (B&C)	Up to 0.1 in	0.006 in	Flatness Block  Oshkosh, WI
Gas Flow	50 sccm to 100 slm	0.38 % of reading + 0.000 22 slm	Fluke molbox/molbloc Calibration System, Mass Flow Controller  Stacy, MN
Liquid Flowmeters	(0.4 to 2) lpm	0.16 % of reading + 0.027 lpm	Comparison to Omega FLR1000 Flowmeter  Oshkosh, WI
Liquid Flow – Syringe	(1 to 1 500) ml/h	0.074 % of reading + 20 µl/h	Syringe Pump, Master Syringe  Oshkosh, WI

## Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Optical Power – Measure <sup>1</sup> (800 to 1 650) nm	(-70 to 20) dBm	0.03 dB/dBm	Agilent 81533B Interface, 81525A Optical Head  Oshkosh, WI Fenton, MO Stacy, MN
Optical Power – Source <sup>1</sup> (820, 1 310, 1 550) nm	(-60 to 0) dB	0.05 dB/dB + 0.05 dB	Agilent 81554SM Laser Source Module, 81533B Interface, 81525A Optical Head, 81655A Laser Module, 81570A Optical Attenuator, 81578A Optical Attenuator  Oshkosh, WI Fenton, MO Stacy, MN
Optical Attenuation – Source <sup>1</sup> (700 to 1 650) nm	(-60 to 0) dB	0.04 dB/dB + 0.04 dB	Agilent 81570A and 81578A Optical Attenuators  Oshkosh, WI Fenton, MO Stacy, MN
Optical Wavelength – Measure <sup>1</sup>	(700 to 1 650) nm	0.05 nm	Agilent 86120B Multi-Wavelength Meter  Oshkosh, WI Fenton, MO Stacy, MN
Gloss Meters <sup>2</sup> 20°, 60°, 85°	(0 to 100) GU	0.73 GU	Standard Gloss Tiles  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

### Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Illuminance – Lux Meters	(180 to 18 000) lux	6.9% of reading + 0.66 Lux	Illuminance Projector, Photometric Calibration System Oshkosh, WI

### Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature – Source	(-95 to 600) °C (600 to 1200) °C	0.03 °C 2.3 °C	Fluke 9011, 9190A Drywell, PRT, Type S Probe  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Infrared Thermometers <sup>1,6</sup>	125 °F 200 °F 400 °F 500 °F 900 °F 932 °F	2.3 °F 2.9 °F 4.4 °F 5.2 °F 8.2 °F 8.5 °F	Hart Scientific 9132 Blackbody $\lambda = (8 \text{ to } 14) \mu\text{m}$ , $\varepsilon = 0.95$  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Surface Probes <sup>1</sup>	(35 to 400) °C	1.3 °C	Hart Scientific 2200 Temperature Controller, Hot-plate  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Temperature – Measure <sup>1</sup>	(-30 to 600) °C (600 to 1 200) °C	0.03 °C 1.7 °C	Hart Scientific 1502 Indicator, PRT, Type S Probe  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Thermo-Hygrometers			Thunder Scientific 2500 Humidity Chamber
Temperature	(0 to 70) °C	0.2 °C	Oshkosh, WI Fenton, MO Stacy, MN Monroe, NC
Humidity	(10 to 98) %RH	0.9 %RH	
System Accuracy Test <sup>1</sup> (SAT)	(0 to 2 200) °F	2.6 °F	Certified Thermocouple  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

### Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature Uniformity Survey (TUS) <sup>1</sup>	(0 to 2 200) °F	4.9 °F	MV 1000 Data Logger or Equivalent  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

### Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Time Interval <sup>1</sup>	(1 to 86 400) s	450 µs	Agilent 53132A Counter and Spectracom 8197B GPS Oscillator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Frequency – Measure <sup>1</sup>	0.1 Hz to 26.5 GHz	7.6 % of reading	Agilent N5531S Measuring Receiver, Spectracom 8197B GPS Oscillator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Source <sup>1</sup>	10 MHz	24 pHz	Spectracom 8197B GPS Oscillator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC
Frequency – Source <sup>1</sup>	0.1 mHz to 40 GHz	4.1 % of reading	Agilent 3325B Function Generator, Agilent E8257D Signal Generator, SRS FS725 Frequency Standard  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC
Tachometers <sup>1</sup> Contact	(1 to 6 500) rpm	0.08 % of reading	King Nutronics 3711-B Tachometer Test Set  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC
Tachometers <sup>1</sup> Non-Contact	(500 to 40 000) rpm	0.08 % of reading	Crestline, OH Tahlequah, OK
Tachometers <sup>1</sup> Non-Contact	(0.01 to 100 000) rpm	0.005 % of reading	Fluke 5522A Multiproduct Calibrator  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK

## DIMENSIONAL MEASUREMENT

### 1 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Linear	Up to 24 in	0.000 9 in	Video Measuring Machine  Oshkosh, WI Fenton, MO Stacy, MN

### 2 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Surface Finish (Ra)	(0.01 to 600) $\mu$ in	2.1 $\mu$ in	Profilometer  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC Crestline, OH Tahlequah, OK
Angular	Up to 180°	0.002 5°	Starrett AVR300  Oshkosh, WI Fenton, MO Stacy, MN Rockford, IL Monroe, NC

### 3 Dimensional


Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Dimensional Inspection <sup>2</sup> Volumetric	Up to (28 x 40 x 24) in	320 $\mu$ in	Coordinate Measuring Machine
Linear	Up to (28 x 40 x 24) in	(38 + 5.2L) $\mu$ in	Oshkosh, WI Stacy, MN



Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2.  $L$  = length in inches,  $D$  = diameter in inches,  $DL$  = diagonal length in inches, GU = Gloss Unit.
3. Where ranges overlap, the uncertainty associated with the higher range begins above the overlapping value.
4. RF/Microwave uncertainties do not include inaccuracies due to sensor mismatch.
5. The CMC for scales and balances is highly dependent upon the resolution of the unit under test. The CMC presented here does not include the resolution of the unit under test. The resolution will be included in the reported measurement uncertainty at the time of calibration.
6. The fixed values presented here are approximate values. Actual calibration values may be used at the time of calibration.
7. This scope is formatted as part of a single document including Certificate of Accreditation No. ACT-1272.



Jason Stine, Vice President

