

**CERTIFICATE OF
 CALIBRATION**



Certificate No. ACT-1272

CERTIFICATION NUMBER	CP134-00040-840	IDENTIFICATION	EL-051A
FOR	Sample Manufacturing 3121 Medalist Drive Oshkosh, WI 54902	SERIAL NUMBER	40
TEST INSTRUMENT	Air Ionizer	PURCHASE ORDER #	
MAKE	Desco	PROCEDURES FOLLOWED	EL-051A rev. 1
MODEL	60505		
CUSTOMER LOCATION	Location1 Location2	STANDARDS USED	
CONDITION RECEIVED	In Tolerance	INSTRUMENT	SERIAL NUMBER
CONDITION RETURNED	In Tolerance	T-0921	16781039.76
CALIBRATED BY	Brian L Gliszinski	T-0953	486
REVIEWED/ISSUED	D.P. 07/29/2025	TRACE NUMBER	NEXT CAL
CALIBRATION LOCATION	FVM	C0151-17646-690	05/31/2025
ENVIRONMENT	69.0°F, 31.0%RH, 28.92inHg	CP065-27852-856	03/31/2026
CALIBRATION DATE	05/14/2025		
RECALIBRATION DUE	05/14/2026		



CALIBRATION RESULTS

* DENOTES "OUT OF TOLERANCE"

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
Positive Decay	(Sec)	(Sec)	(Sec)	(Sec)	(Sec)	(Sec)
Test						
Run 1	0.0	0.0	2.0	1.9	1.9	6.0E-2
Run 2	0.0	0.0	2.0	1.7	1.7	6.0E-2
Run 3	0.0	0.0	2.0	1.7	1.7	6.0E-2
Run 4	0.0	0.0	2.0	1.8	1.8	6.0E-2
Run 5	0.0	0.0	2.0	1.7	1.7	6.0E-2
Run 6	0.0	0.0	2.0	1.6	1.6	6.0E-2
MIN	0.0	0.0	2.0	1.6	1.6	6.0E-2
MAX	0.0	0.0	2.0	1.9	1.9	6.0E-2
AVG	0.00	0.00	2.00	1.73	1.73	6.0E-3
Negative Decay	(Sec)	(Sec)	(Sec)	(Sec)	(Sec)	(Sec)
Test						
Run 1	0.0	0.0	2.0	1.7	1.7	6.0E-2
Run 2	0.0	0.0	2.0	1.7	1.7	6.0E-2
Run 3	0.0	0.0	2.0	1.7	1.7	6.0E-2
Run 4	0.0	0.0	2.0	1.7	1.7	6.0E-2
Run 5	0.0	0.0	2.0	1.9	1.9	6.0E-2
Run 6	0.0	0.0	2.0	1.7	1.7	6.0E-2
MIN	0.0	0.0	2.0	1.7	1.7	6.0E-2
MAX	0.0	0.0	2.0	1.9	1.9	6.0E-2
AVG	0.00	0.00	2.00	1.73	1.73	6.0E-3
Ion Balance	(V)	(V)	(V)	(V)	(V)	(V)
	0	-10	10	-8	-8	6.0E-1
Alarm						

Fox Valley Metrology

3114 Medalist Drive

Oshkosh, WI 54902

(920) 426-5894 • Fax (920) 426-8120

<https://www.FoxValleyMetrology.com>

CERTIFICATE OF CALIBRATION



Certificate No. ACT-1272

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
alarm and light with front grills shorted.	Pass/Fail	Pass/Fail	Pass/Fail	Pass	Pass	

COMMENTS

Example certificate, actual results and uncertainties will be reported at the time of calibration.
Nominal and limits may vary based on actual make/model.

- This certificate shall not be altered in any form or reproduced, except in full, without prior written approval from originating lab. These results relate only to the item(s) calibrated. Form Revision 10: 06/04/2024
- Total expanded measurement uncertainties expressed are based on a confidence level of 95%; coverage factor of (k=2). The statement of compliance in this certificate was issued without taking the uncertainty of measurement into consideration. The customer shall assess the results and uncertainty when determining if the results meet their needs. (This is considered "shared responsibility.") Uncertainties expressed in nominal units.
- The calibrations within the certificate/report are traceable through NIST or another National Metrology Institute to the International System of Units (SI). Calibration was completed in accordance with ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. Other standards listed upon request.