

**CERTIFICATE OF  
 CALIBRATION**



Certificate No. ACT-1272

<p><b>CERTIFICATION NUMBER</b> CP134-00374-840</p> <p><b>FOR</b> Sample Manufacturing          3121 Medalist Drive          Oshkosh, WI 54902</p> <p><b>TEST INSTRUMENT</b> Analyzer          Harmonic</p> <p><b>MAKE</b> Fluke</p> <p><b>MODEL</b> 43B</p> <p><b>CUSTOMER LOCATION</b> Location1          Location2</p> <p><b>CONDITION RECEIVED</b> In Tolerance</p> <p><b>CONDITION RETURNED</b> In Tolerance</p> <p><b>CALIBRATED BY</b> Brian L Gliszinski</p> <p><b>REVIEWED/ISSUED</b> D.P. 07/29/2025</p> <p><b>CALIBRATION LOCATION</b> FVM</p> <p><b>ENVIRONMENT</b> 69.0°F, 31.0%RH, 28.92inHg</p> <p><b>CALIBRATION DATE</b> 05/14/2025</p> <p><b>RECALIBRATION DUE</b> 05/14/2026</p>	<p><b>IDENTIFICATION</b> EL-023A</p> <p><b>SERIAL NUMBER</b> 374</p> <p><b>PURCHASE ORDER #</b></p> <p><b>PROCEDURES FOLLOWED</b>          EL-023A rev. 2</p> <p><b>STANDARDS USED</b></p> <table border="1"> <thead> <tr> <th>INSTRUMENT</th> <th>SERIAL NUMBER</th> <th>TRACE NUMBER</th> <th>NEXT CAL</th> </tr> </thead> <tbody> <tr> <td>COIL-001</td> <td>20083512</td> <td>CL088-28770-531</td> <td>07/31/2025</td> </tr> <tr> <td>FVS-845</td> <td>FVS-845</td> <td>C0191-37540-690</td> <td>07/31/2025</td> </tr> <tr> <td>FVS-A05</td> <td>6375609</td> <td>CP042-30961-087</td> <td>07/31/2025</td> </tr> </tbody> </table>	INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL	COIL-001	20083512	CL088-28770-531	07/31/2025	FVS-845	FVS-845	C0191-37540-690	07/31/2025	FVS-A05	6375609	CP042-30961-087	07/31/2025
INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL														
COIL-001	20083512	CL088-28770-531	07/31/2025														
FVS-845	FVS-845	C0191-37540-690	07/31/2025														
FVS-A05	6375609	CP042-30961-087	07/31/2025														



**CALIBRATION RESULTS**

\* DENOTES "OUT OF TOLERANCE"

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
INSTALLED INSTRUMENT FIRMWARE						
Firmware Version: V2.03						
BACKLIGHT POWER SAVE						
DISPLAY TESTS						
Maximum Darkness Result of Operator Evaluation						
Default Brightness Result of Operator Evaluation						
Maximum Brightness						

**Reviewed**

**CERTIFICATE OF  
 CALIBRATION**



Certificate No. ACT-1272

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
Result of Operator Evaluation						
FREQUENCY RESPONSE, Input 1						
0.00 dB @ 20 MHz	0	-3	3	0	0	6.0E-1
FREQUENCY MEASUREMENT ACCURACY						
Input 1						
1.00 MHz @ 1.2 Vpp	1.00	0.97	1.02	1.00	1.00	6.0E-3
10.0 MHz @ 1.2 Vpp	10.0	9.7	10.3	10.0	10.0	6.0E-2
40.0 MHz @ 2.4 Vpp	40.0	38.8	41.2	40.0	40.0	6.0E-2
Input 2						
15.0 kHz @ 0.6 Vpp	15.0	14.7	15.3	15.0	15.0	6.0E-2
TRIGGER SENSITIVITY, Input 1						
5 MHz						
25 MHz						
40 MHz						
TRIGGER SENSITIVITY, Input 2						
5 kHz						
20 kHz						
TRIGGER LEVEL and SLOPE TEST						
Input 1						
2 V @ (+)SLOPE	2	2	3	2	2	6.0E-1
2 V @ (-)SLOPE	2	2	3	2	2	6.0E-1
Input 2						
2 V @ (+)SLOPE	2	2	3	2	2	6.0E-1
2 V @ (-)SLOPE	2	2	3	2	2	6.0E-1

**CERTIFICATE OF  
 CALIBRATION**



Certificate No. ACT-1272

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
DC VOLTAGE (Input 1)						
5 mV/div 15.0 mV	15.0	14.4	15.6	15.0	15.0	6.0E-2
10 mV/div 30.0 mV	30.0	29.4	30.6	30.0	30.0	6.0E-2
20 mV/div 60.0 mV	60.0	59.2	60.8	60.0	60.0	6.0E-2
50 mV/div 150.0 mV	150.0	148.8	151.3	150.0	150.0	6.0E-2
100 mV/div 300.0 mV	300.0	298.0	302.0	300.0	300.0	6.0E-2
200 mV/div 0.0 mV	0.0	-0.5	0.5	0.0	0.0	6.0E-2
500.0 mV	500.0	497.0	503.0	500.0	500.0	6.0E-2
-500.0 mV	-500.0	-503.0	-497.0	-500.0	-500.0	6.0E-2
500 mV/div 1.500 V	1.500	1.488	1.513	1.500	1.500	6.0E-4
1 V/div 3.000 V	3.000	2.980	3.020	3.000	3.000	6.0E-4
2 V/div 0.000 V	0.000	-0.005	0.005	0.000	0.000	6.0E-4
5.000 V	5.000	4.970	5.030	5.000	5.000	6.0E-4
-5.000 V	-5.000	-5.030	-4.970	-5.000	-5.000	6.0E-4
5 V/div 15.00 V	15.00	14.88	15.13	15.00	15.00	6.0E-3
10 V/div 30.00 V	30.00	29.80	30.20	30.00	30.00	6.0E-3
20 V/div 0.00 V	0.00	-0.05	0.05	0.00	0.00	6.0E-3
50.00 V	50.00	49.70	50.30	50.00	50.00	6.0E-3
-50.00 V	-50.00	-50.30	-49.70	-50.00	-50.00	6.0E-3
50 V/div 150.0 V	150.0	148.8	151.3	150.0	150.0	6.0E-2
100 V/div 300.0 V	300.0	298.0	302.0	300.0	300.0	6.0E-2

**CERTIFICATE OF  
 CALIBRATION**



Certificate No. ACT-1272

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
200 V/div						
0.0 V	0.0	-0.5	0.5	0.0	0.0	6.0E-2
500.0 V	500.0	497.0	503.0	500.0	500.0	6.0E-2
-500.0 V	-500.0	-503.0	-497.0	-500.0	-500.0	6.0E-2
500 V/div						
0 V	0	-5	5	0	0	6.0E-1
600 V	600	592	608	600	600	6.0E-1
-600 V	-600	-608	-592	-600	-600	6.0E-1
DC CURRENT (Input 2)						
5 A/div						
15.00 A	15.00	14.88	15.13	15.00	15.00	6.0E-3
10 A/div						
30.00 A	30.00	29.80	30.20	30.00	30.00	6.0E-3
20 A/div						
60.00 A	60.00	59.65	60.35	60.00	60.00	6.0E-3
50 A/div						
150.0 A	150.0	148.8	151.3	150.0	150.0	6.0E-2
100 A/div						
300.0 A	300.0	298.0	302.0	300.0	300.0	6.0E-2
200 A/div						
0.0 A	0.0	-0.5	0.5	0.0	0.0	6.0E-2
500.0 A	500.0	497.0	503.0	500.0	500.0	6.0E-2
-500.0 A	-500.0	-503.0	-497.0	-500.0	-500.0	6.0E-2
500 A/div						
1.500 kA	1.500	1.488	1.513	1.500	1.500	6.0E-4
1 kA/div						
3.000 kA	3.000	2.980	3.020	3.000	3.000	6.0E-4
2 kA/div						
0.000 kA	0.000	-0.005	0.005	0.000	0.000	6.0E-4
5.000 kA	5.000	4.970	5.030	5.000	5.000	6.0E-4
-5.000 kA	-5.000	-5.030	-4.970	-5.000	-5.000	6.0E-4
5 kA/div						
15.00 kA	15.00	14.88	15.13	15.00	15.00	6.0E-3
10 kA/div						
30.00 kA	30.00	29.80	30.20	30.00	30.00	6.0E-3
20 kA/div						

**CERTIFICATE OF  
 CALIBRATION**



Certificate No. ACT-1272

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
0.00 kA	0.00	-0.05	0.05	0.00	0.00	6.0E-3
50.00 kA	50.00	49.70	50.30	50.00	50.00	6.0E-3
-50.00 kA	-50.00	-50.30	-49.70	-50.00	-50.00	6.0E-3
50 kA/div						
150.0 kA	150.0	148.8	151.3	150.0	150.0	6.0E-2
100 kA/div						
300.0 kA	300.0	298.0	302.0	300.0	300.0	6.0E-2
200 kA/div						
0.0 kA	0.0	-0.5	0.5	0.0	0.0	6.0E-2
500.0 kA	500.0	497.0	503.0	500.0	500.0	6.0E-2
-500.0 kA	-500.0	-503.0	-497.0	-500.0	-500.0	6.0E-2
500 kA/div						
0 kA	0	0	0	0	0	6.0E-1
600 kA	600	592	608	600	600	6.0E-1
-600 kA	-600	-608	-592	-600	-600	6.0E-1
AC VOLTAGE (Input 1)						
200 mV/div						
500.0 mV @ 60 Hz	500.0	494.0	506.0	500.0	500.0	6.0E-2
500.0 mV @ 20 kHz	500.0	486.0	514.0	500.0	500.0	6.0E-2
2 V/div						
5.000 V @ 60 Hz	5.000	4.940	5.060	5.000	5.000	6.0E-4
5.000 V @ 20 kHz	5.000	4.860	5.140	5.000	5.000	6.0E-4
20 V/div						
50.00 V @ 60 Hz	50.00	49.40	50.60	50.00	50.00	6.0E-3
50.00 V @ 20 kHz	50.00	48.60	51.40	50.00	50.00	6.0E-3
200 V/div						
500.0 V @ 60 Hz	500.0	494.0	506.0	500.0	500.0	6.0E-2
500.0 V @ 10 kHz	500.0	486.0	514.0	500.0	500.0	6.0E-2
500 V/div						
600 V @ 60 Hz	600	584	616	600	600	6.0E-1
600 V @ 10 kHz	600	570	630	600	600	6.0E-1
AC CURRENT (Input 2)						
200 A/div						

**CERTIFICATE OF  
 CALIBRATION**



Certificate No. ACT-1272

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
500.0 A @ 60 Hz	500.0	494.0	506.0	500.0	500.0	6.0E-2
2 kA/div						
5.000 kA @ 60 Hz	5.000	4.940	5.060	5.000	5.000	6.0E-4
20 kA/div						
50.00 kA @ 60 Hz	50.00	49.40	50.60	50.00	50.00	6.0E-3
200 kA/div						
500.0 kA @ 60 Hz	500.0	494.0	506.0	500.0	500.0	6.0E-2
500 kA/div						
600 kA @ 60 Hz	600	584	616	600	600	6.0E-1
AC VOLTAGE (Input 1, AC Coupling)						
200 mV/div						
500.0 mV @ 10 Hz	500.0	344.0	500.0	500.0	500.0	6.0E-2
500.0 mV @ 33 Hz	500.0	469.0	500.0	500.0	500.0	6.0E-2
500.0 mV @ 60 Hz	500.0	486.5	500.0	500.0	500.0	6.0E-2
AC CURRENT (Input 2, AC Coupling)						
200 A/div						
500.0 A @ 10 Hz	500.0	344.0	500.0	500.0	500.0	6.0E-2
500.0 A @ 33 Hz	500.0	469.0	500.0	500.0	500.0	6.0E-2
500.0 A @ 60 Hz	500.0	486.0	500.0	500.0	500.0	6.0E-2
AC VOLTAGE (Input 1, Volts Peak)						
5 Vpp	5	5	6	5	5	6.0E-1
AC CURRENT (Input 2, Amps Peak)						
5 kApp	5	5	6	5	5	6.0E-1
PHASE MEASUREMENT (Input 1 to Input 2)						
0 °	0	-2	2	0	0	6.0E-1

**CERTIFICATE OF  
 CALIBRATION**



Certificate No. ACT-1272

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
<b>RESISTANCE</b>						
0.0 Ohm	0.0	-0.5	0.5	0.0	0.0	6.0E-2
400.0 Ohm	400.0	397.1	402.9	400.0	400.0	6.0E-2
4.000 kOhm	4.000	3.971	4.029	4.000	4.000	6.0E-4
40.00 kOhm	40.00	39.71	40.29	40.00	40.00	6.0E-3
400.0 kOhm	400.0	397.1	402.9	400.0	400.0	6.0E-2
4.000 MOhm	4.000	3.971	4.029	4.000	4.000	6.0E-4
30.00 MOhm	30.00	29.77	30.23	30.00	30.00	6.0E-3
<b>CONTINUITY</b>						
Beeper Audible						
Beeper NOT Audible						
<b>DIODE TEST</b>						
Current						
500.0 mV	500.0	425.0	575.0	500.0	500.0	6.0E-2
Voltage						
1.000 V	1.000	0.975	1.025	1.000	1.000	6.0E-4
<b>CAPACITANCE</b>						
50 nF Range						
40.00 nF	40.00	39.10	40.90	40.00	40.00	6.0E-3
500 nF Range						
300.0 nF	300.0	293.0	307.0	300.0	300.0	6.0E-2
5 µF Range						
3.000 µF	3.000	2.930	3.070	3.000	3.000	6.0E-4
50 µF Range						
30.00 µF	30.00	29.30	30.70	30.00	30.00	6.0E-3
500 µF Range						
300.0 µF	300.0	293.0	307.0	300.0	300.0	6.0E-2
50 nF Range						
0.00 nF	0.00	0.00	0.10	0.00	0.00	6.0E-3
<b>INRUSH CURRENT MODE</b>						
Cursor Position 1						
1.50 kA	1.50	1.38	1.63	1.50	1.50	6.0E-3
Cursor Position 2						

**CERTIFICATE OF  
 CALIBRATION**



Certificate No. ACT-1272

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
1.50 kA	1.50	1.38	1.63	1.50	1.50	6.0E-3
SAGS & SWELLS						
Volts						
5.00 V	5.00	4.80	5.20	5.00	5.00	6.0E-3
Amps						
5.00 V	5.00	4.80	5.20	5.00	5.00	6.0E-3
HARMONICS						
Result of Operator Evaluation						
Result of Operator Evaluation						
VOLTS/AMPS/HERTZ Z MODE						
Volts						
0.000 V	0.000	-0.010	0.010	0.000	0.000	6.0E-4
5.500 V @ 70 Hz	5.500	5.340	5.660	5.500	5.500	6.0E-4
3.500 V @ 15 kHz	3.500	3.300	3.700	3.500	3.500	6.0E-4
Amps						
0.000 A	0.000	-0.100	0.100	0.000	0.000	6.0E-4
5.500 kA @ 70 Hz	5.500	5.340	5.660	5.500	5.500	6.0E-4
4.500 kA @ 70 Hz	4.500	4.445	4.555	4.500	4.500	6.0E-4
3.500 kA @ 15 kHz	3.500	3.300	3.700	3.500	3.500	6.0E-4
Hertz						
70.0 Hz @ 5.5 V	70.0	69.4	70.6	70.0	70.0	6.0E-2
70.0 Hz @ 4.5 kA	70.0	69.4	70.6	70.0	70.0	6.0E-2
POWER TEST						
0 V Applied						
0.00 W	0.00	-0.04	0.04	0.00	0.00	6.0E-3
0.00 VA	0.00	-0.04	0.04	0.00	0.00	6.0E-3
0.00 VAR_	0.00	-0.04	0.04	0.00	0.00	6.0E-3
4.472 V Applied						
20.0 kW @ 60 Hz	20.0	19.4	20.6	20.0	20.0	6.0E-2
20.0 kVA @ 60 Hz	20.0	19.4	20.6	20.0	20.0	6.0E-2

**CERTIFICATE OF  
 CALIBRATION**



Certificate No. ACT-1272

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
0.00 kVAR_ @ 60 Hz	0.00	-0.40	0.40	0.00	0.00	6.0E-3
1.00 PF @ 60 Hz	1.00	0.96	1.00	1.00	1.00	6.0E-3
1.00 DPF @ 60 Hz	1.00	0.97	1.00	1.00	1.00	6.0E-3
60.0 Hz @ 4.472 V	60.0	59.5	60.5	60.0	60.0	6.0E-2
5.916 V Applied						
35.0 kW @ 60 Hz	35.0	34.3	35.7	35.0	35.0	6.0E-2
35.0 kVA @ 60 Hz	35.0	34.3	35.7	35.0	35.0	6.0E-2
0 VAR_ @ 60 Hz	0	-4	4	0	0	6.0E-1
1.00 PF	1.00	0.96	1.00	1.00	1.00	6.0E-3
1.00 DPF	1.00	0.97	1.00	1.00	1.00	6.0E-3
60.0 Hz @ 5.916 V	60.0	59.5	60.5	60.0	60.0	6.0E-2
TRANSIENT MODE						
No Transients						
Captured						
Transients						
Captured						

**COMMENTS**

Example certificate, actual results and uncertainties will be reported at the time of calibration.  
 Nominal and limits may vary based on actual make/model.

- This certificate shall not be altered in any form or reproduced, except in full, without prior written approval from originating lab. These results relate only to the item(s) calibrated. Form Revision 10: 06/04/2024
- Total expanded measurement uncertainties expressed are based on a confidence level of 95%; coverage factor of (k=2). The statement of compliance in this certificate was issued without taking the uncertainty of measurement into consideration. The customer shall assess the results and uncertainty when determining if the results meet their needs. (This is considered "shared responsibility.") Uncertainties expressed in nominal units.
- The calibrations within the certificate/report are traceable through NIST or another National Metrology Institute to the International System of Units (SI). Calibration was completed in accordance with ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. Other standards listed upon request.