

**CERTIFICATE OF  
 CALIBRATION**



Certificate No. ACT-1272

<p><b>CERTIFICATION NUMBER</b> CP134-00196-840</p> <p><b>FOR</b> Sample Manufacturing          3121 Medalist Drive          Oshkosh, WI 54902</p> <p><b>TEST INSTRUMENT</b> Calipers          Digital</p> <p><b>MAKE</b> Mitutoyo</p> <p><b>MODEL</b> 500-196-30</p> <p><b>RANGE</b> (0 to 6) in</p> <p><b>CUSTOMER LOCATION</b> Location1          Location2</p> <p><b>CONDITION RECEIVED</b> In Tolerance</p> <p><b>CONDITION RETURNED</b> In Tolerance</p> <p><b>CALIBRATED BY</b> Jim Braun</p> <p><b>REVIEWED/ISSUED</b> D.P. 07/29/2025</p> <p><b>CALIBRATION LOCATION</b> FVM</p> <p><b>ENVIRONMENT</b> 69.0°F, 31.0%RH, 28.92inHg</p> <p><b>CALIBRATION DATE</b> 05/14/2025</p> <p><b>RECALIBRATION DUE</b> 05/14/2026</p>	<p><b>IDENTIFICATION</b> HT-004A</p> <p><b>SERIAL NUMBER</b> 196</p> <p><b>PURCHASE ORDER #</b></p> <p><b>PROCEDURES FOLLOWED</b>          HT-004A rev. 1</p> <p><b>STANDARDS USED</b></p> <table border="1"> <thead> <tr> <th>INSTRUMENT</th> <th>SERIAL NUMBER</th> <th>TRACE NUMBER</th> <th>NEXT CAL</th> </tr> </thead> <tbody> <tr> <td>FVM-142</td> <td>8PCS-130013</td> <td>C0227-49403-662</td> <td>08/31/2025</td> </tr> <tr> <td>FVR-005-183</td> <td></td> <td>C0169-52119-590</td> <td>06/30/2025</td> </tr> <tr> <td>FVR-005-207</td> <td></td> <td>C0170-40594-388</td> <td>06/30/2025</td> </tr> <tr> <td>FVR-059</td> <td>80196</td> <td>C0229-26351-590</td> <td>08/31/2025</td> </tr> <tr> <td>FVR-119B</td> <td>16782367.21</td> <td>C0213-33882-690</td> <td>07/31/2025</td> </tr> <tr> <td>FVR-248</td> <td></td> <td>CP014-37602-379</td> <td>01/31/2026</td> </tr> </tbody> </table>	INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL	FVM-142	8PCS-130013	C0227-49403-662	08/31/2025	FVR-005-183		C0169-52119-590	06/30/2025	FVR-005-207		C0170-40594-388	06/30/2025	FVR-059	80196	C0229-26351-590	08/31/2025	FVR-119B	16782367.21	C0213-33882-690	07/31/2025	FVR-248		CP014-37602-379	01/31/2026
INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL																										
FVM-142	8PCS-130013	C0227-49403-662	08/31/2025																										
FVR-005-183		C0169-52119-590	06/30/2025																										
FVR-005-207		C0170-40594-388	06/30/2025																										
FVR-059	80196	C0229-26351-590	08/31/2025																										
FVR-119B	16782367.21	C0213-33882-690	07/31/2025																										
FVR-248		CP014-37602-379	01/31/2026																										



**CALIBRATION RESULTS**

\* DENOTES "OUT OF TOLERANCE"

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
OD Jaws	(in)	(in)	(in)	(in)	(in)	(in)
	0.3700	0.3690	0.3710	0.3700	0.3700	6.0E-5
	1.0000	0.9990	1.0010	1.0000	1.0000	6.0E-5
	1.6000	1.5990	1.6010	1.6000	1.6000	6.0E-5
	3.0000	2.9990	3.0010	3.0000	3.0000	6.0E-5
	3.7500	3.7490	3.7510	3.7500	3.7500	6.0E-5
	5.0000	4.9990	5.0010	5.0000	5.0000	6.0E-5
ID Jaws	(in)	(in)	(in)	(in)	(in)	(in)
	0.7000	0.6990	0.7010	0.7000	0.7000	6.0E-5
Step	(in)	(in)	(in)	(in)	(in)	(in)
	1.0000	0.9990	1.0010	1.0000	1.0000	6.0E-5
Depth	(in)	(in)	(in)	(in)	(in)	(in)
	1.0000	0.9990	1.0010	1.0000	1.0000	6.0E-5

**COMMENTS**

Example certificate, actual results and uncertainties will be reported at the time of calibration.  
 Nominal and limits may vary based on actual make/model.

## Fox Valley Metrology

3114 Medalist Drive

Oshkosh, WI 54902

(920) 426-5894 • Fax (920) 426-8120

<https://www.FoxValleyMetrology.com>

## CERTIFICATE OF CALIBRATION



Certificate No. ACT-1272

- 
- This certificate shall not be altered in any form or reproduced, except in full, without prior written approval from originating lab. These results relate only to the item(s) calibrated. Form Revision 10: 06/04/2024
  - Total expanded measurement uncertainties expressed are based on a confidence level of 95%; coverage factor of (k=2). The statement of compliance in this certificate was issued without taking the uncertainty of measurement into consideration. The customer shall assess the results and uncertainty when determining if the results meet their needs. (This is considered "shared responsibility.") Uncertainties expressed in nominal units.
  - The calibrations within the certificate/report are traceable through NIST or another National Metrology Institute to the International System of Units (SI). Calibration was completed in accordance with ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. Other standards listed upon request.