

Fox Valley Metrology

3114 Medalist Drive

Oshkosh, WI 54902

(920) 426-5894 • Fax (920) 426-8120

<https://www.FoxValleyMetrology.com>

CERTIFICATE OF CALIBRATION



Certificate No. ACT-1272

CERTIFICATION NUMBER	CP134-00313-840	IDENTIFICATION	EL-029A
FOR	Sample Manufacturing 3121 Medalist Drive Oshkosh, WI 54902	SERIAL NUMBER	313
TEST INSTRUMENT	Current Probe	PURCHASE ORDER #	
MAKE	Tektronix	PROCEDURES FOLLOWED	EL-029A rev. 1
MODEL	P6021A		
CUSTOMER LOCATION	Location1 Location2	STANDARDS USED	
CONDITION RECEIVED	In Tolerance	INSTRUMENT	SERIAL NUMBER
CONDITION RETURNED	In Tolerance	T-0834	3968901
CALIBRATED BY	Brian L Gliszinski	T-0952	C030968
REVIEWED/ISSUED	D.P. 07/29/2025	T-1017	
CALIBRATION LOCATION	FVM	TRACE NUMBER	NEXT CAL
ENVIRONMENT	69.0°F, 31.0%RH, 28.92inHg	CP091-33194-762	07/31/2025
CALIBRATION DATE	05/14/2025	C0171-64901-838	07/31/2025
RECALIBRATION DUE	05/14/2026	CP052-49561-802	07/31/2025

CALIBRATION RESULTS

* DENOTES "OUT OF TOLERANCE"

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
Midband Accuracy	%	%	%	%	%	%
2mA/mV Range	0.0	-3.0	3.0	0.1	0.1	6.0E-2
10mA/mV Range	0.0	-3.0	3.0	0.0	0.0	6.0E-2
Rise Time						
2mA/mV Range	<5.8ns	<5.8ns	<5.8ns	Pass	Pass	
10mA/mV Range	<5.8ns	<5.8ns	<5.8ns	Pass	Pass	
Calculated Bandwidth High Limit						
2mA/mV Range	>60MHz	>60MHz	>60MHz	Pass	Pass	
10mA/mV Range	>60MHz	>60MHz	>60MHz	Pass	Pass	
Low Limit						
2mA/mV Range	<1kHz	<1kHz	<1kHz	Pass	Pass	
10mA/mV Range	<300Hz	<300Hz	<300Hz	Pass	Pass	
Time Constant						
2mA/mV Range	>159us	>159us	>159us	Pass	Pass	
10mA/mV Range	>530us	>530us	>530us	Pass	Pass	

COMMENTS

Example certificate, actual results and uncertainties will be reported at the time of calibration.
Nominal and limits may vary based on actual make/model.

Fox Valley Metrology

3114 Medalist Drive

Oshkosh, WI 54902

(920) 426-5894 • Fax (920) 426-8120

<https://www.FoxValleyMetrology.com>

CERTIFICATE OF CALIBRATION



Certificate No. ACT-1272

-
- This certificate shall not be altered in any form or reproduced, except in full, without prior written approval from originating lab. These results relate only to the item(s) calibrated. Form Revision 10: 06/04/2024
 - Total expanded measurement uncertainties expressed are based on a confidence level of 95%; coverage factor of (k=2). The statement of compliance in this certificate was issued without taking the uncertainty of measurement into consideration. The customer shall assess the results and uncertainty when determining if the results meet their needs. (This is considered "shared responsibility.") Uncertainties expressed in nominal units.
 - The calibrations within the certificate/report are traceable through NIST or another National Metrology Institute to the International System of Units (SI). Calibration was completed in accordance with ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. Other standards listed upon request.
-