

**CERTIFICATE OF
 CALIBRATION**



Certificate No. ACT-1272

CERTIFICATION NUMBER	CP134-00104-840	IDENTIFICATION	EL-017A
FOR	Sample Manufacturing 3121 Medalist Drive Oshkosh, WI 54902	SERIAL NUMBER	104
TEST INSTRUMENT	Frequency Counter	PURCHASE ORDER #	
MAKE	Keysight	PROCEDURES FOLLOWED	EL-017A rev. 1
MODEL	53131A	STANDARDS USED	
CUSTOMER LOCATION	Location1 Location2	INSTRUMENT	SERIAL NUMBER
CONDITION RECEIVED	In Tolerance	T-0753	2884903
CONDITION RETURNED	In Tolerance	T-0884	3546A02489
CALIBRATED BY	Jeremy Armstrong	T-0968	C8922
REVIEWED/ISSUED	D.P. 07/29/2025		
CALIBRATION LOCATION	FVM	TRACE NUMBER	NEXT CAL
ENVIRONMENT	69.0°F, 31.0%RH, 28.92inHg	CP105-41963-762	04/30/2026
CALIBRATION DATE	05/14/2025	C0250-41429-802	09/30/2025
RECALIBRATION DUE	05/14/2026	C0322-55686-697	11/30/2025



CALIBRATION RESULTS

* DENOTES "OUT OF TOLERANCE"

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
Self Test	Pass/Fail			Pass	Pass	
Time Interval	(uS) 0.1500	(uS) 0.1460	(uS) 0.1540	(uS) 0.1536	(uS) 0.1536	(uS) 6.0E-5
Trigger Level	(V)	(V)	(V)	(V)	(V)	(V)
Channel 1	0.000	-0.015	0.015	0.005	0.005	6.0E-4
Channel 2	0.000	-0.015	0.015	0.006	0.006	6.0E-4
Peak Volts	(V)	(V)	(V)	(V)	(V)	(V)
Channel 2 (-)	-1.00	-1.12	-0.88	-0.95	-0.95	6.0E-3
Channel 2 (+)	1.00	0.88	1.12	0.97	0.97	6.0E-3
Channel 1 (-)	-1.00	-1.12	-0.88	-0.96	-0.96	6.0E-3
Channel 1 (+)	1.00	0.88	1.12	0.99	0.99	6.0E-3
Channel 1 Accuracy	(Hz)	(Hz)	(Hz)	(Hz)	(Hz)	(Hz)
	1000000.0000	999999.9990	1000000.0010	1000000.0001	1000000.0001	6.0E-5
	(MHz)	(MHz)	(MHz)	(MHz)	(MHz)	(MHz)
	100.000000000	99.999999930	100.000000070	100.000000000	100.000000000	6.0E-10
	200.000000000	199.999999870	200.000000130	200.000000000	200.000000000	6.0E-10
	225.000000000	224.999999850	225.000000150	225.000000001	225.000000001	6.0E-10
Channel 1 Sensitivity	(dBm)	(dBm)	(dBm)	(dBm)	(dBm)	(dBm)

Fox Valley Metrology

3114 Medalist Drive

Oshkosh, WI 54902

(920) 426-5894 • Fax (920) 426-8120

<https://www.FoxValleyMetrology.com>

CERTIFICATE OF CALIBRATION



Certificate No. ACT-1272

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
100kHz to 100MHz	<-21.0	<-21.0	<-21.0	Pass	Pass	
100MHz to 200MHz	<-17.5	<-17.5	<-17.5	Pass	Pass	
200MHz to 225MHz	<-15.0	<-15.0	<-15.0	Pass	Pass	
Channel 2 Accuracy	(Hz)	(Hz)	(Hz)	(Hz)	(Hz)	(Hz)
	1000000.00000	999999.99900	1000000.00100	1000000.00001	1000000.00001	6.0E-6
	(MHz)	(MHz)	(MHz)	(MHz)	(MHz)	(MHz)
	100.000000000	99.999999930	100.000000070	100.000000000	100.000000000	6.0E-10
	200.000000000	199.999999870	200.000000130	200.000000000	200.000000000	6.0E-10
	225.000000000	224.999999850	225.000000150	225.000000000	225.000000000	6.0E-10
Channel 2 Sensitivity	(dBm)	(dBm)	(dBm)	(dBm)	(dBm)	(dBm)
100kHz to 100MHz	<-21.0	<-21.0	<-21.0	Pass	Pass	
100MHz to 200MHz	<-17.5	<-17.5	<-17.5	Pass	Pass	
200MHz to 225MHz	<-15.0	<-15.0	<-15.0	Pass	Pass	
Time Base	M(Hz)	M(Hz)	M(Hz)	M(Hz)	M(Hz)	
	10.000000	9.999931	10.000069	10.000000	10.000000	6.0E-7
Termination Check	(ck)	(ck)	(ck)	(ck)	(ck)	(ck)
Ch1, 50ohm	Pass/Fail			Pass	Pass	
Ch2, 50ohm	Pass/Fail			Pass	Pass	
Ch1, 1Mohm	Pass/Fail			Pass	Pass	
Ch2, 1Mohm	Pass/Fail			Pass	Pass	
10MHz External Timebase Input	(ck)	(ck)	(ck)	(ck)	(ck)	(ck)
	Pass/Fail			Pass	Pass	
10MHz External Timebase Input	(ck)	(ck)	(ck)	(ck)	(ck)	(ck)
				Pass	Pass	
				Pass	Pass	
				Pass	Pass	
				Pass	Pass	

COMMENTS

Example certificate, actual results and uncertainties will be reported at the time of calibration. Nominal and limits may vary based on actual make/model.

Fox Valley Metrology

3114 Medalist Drive

Oshkosh, WI 54902

(920) 426-5894 • Fax (920) 426-8120

<https://www.FoxValleyMetrology.com>

CERTIFICATE OF CALIBRATION



-
- This certificate shall not be altered in any form or reproduced, except in full, without prior written approval from originating lab. These results relate only to the item(s) calibrated. Form Revision 10: 06/04/2024
 - Total expanded measurement uncertainties expressed are based on a confidence level of 95%; coverage factor of (k=2). The statement of compliance in this certificate was issued without taking the uncertainty of measurement into consideration. The customer shall assess the results and uncertainty when determining if the results meet their needs. (This is considered "shared responsibility.") Uncertainties expressed in nominal units.
 - The calibrations within the certificate/report are traceable through NIST or another National Metrology Institute to the International System of Units (SI). Calibration was completed in accordance with ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. Other standards listed upon request.
-