

**CERTIFICATE OF
 CALIBRATION**



Certificate No. ACT-1272

<p>CERTIFICATION NUMBER CP134-00425-845</p> <p>FOR Sample Manufacturing 3121 Medalist Drive Oshkosh, WI 54902</p> <p>TEST INSTRUMENT Indicators Test</p> <p>MAKE Interapid</p> <p>MODEL 312B-1</p> <p>RANGE (0 to 0.060) in</p> <p>CUSTOMER LOCATION Location1 Location2</p> <p>CONDITION RECEIVED In Tolerance</p> <p>CONDITION RETURNED In Tolerance</p> <p>CALIBRATED BY Jacob Nachtrab</p> <p>REVIEWED/ISSUED D.P. 07/29/2025</p> <p>CALIBRATION LOCATION FVM</p> <p>ENVIRONMENT 69.0°F, 31.0%RH, 28.92inHg</p> <p>CALIBRATION DATE 05/14/2025</p> <p>RECALIBRATION DUE 05/14/2026</p>	<p>IDENTIFICATION HT-016A</p> <p>SERIAL NUMBER 425</p> <p>PURCHASE ORDER #</p> <p>PROCEDURES FOLLOWED HT-016A rev. 1</p> <p>STANDARDS USED</p> <table border="1"> <thead> <tr> <th>INSTRUMENT</th> <th>SERIAL NUMBER</th> <th>TRACE NUMBER</th> <th>NEXT CAL</th> </tr> </thead> <tbody> <tr> <td>STL-397</td> <td>63311376</td> <td>CO218-30712-459</td> <td>08/31/2025</td> </tr> <tr> <td>STL-697</td> <td>N/A</td> <td>CP132-54826-376</td> <td>05/31/2026</td> </tr> </tbody> </table>	INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL	STL-397	63311376	CO218-30712-459	08/31/2025	STL-697	N/A	CP132-54826-376	05/31/2026
INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL										
STL-397	63311376	CO218-30712-459	08/31/2025										
STL-697	N/A	CP132-54826-376	05/31/2026										



CALIBRATION RESULTS

* DENOTES "OUT OF TOLERANCE"

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
Clockwise	(in)	(in)	(in)	(in)	(in)	(in)
	0.0005	0.0000	0.0010	0.0005	0.0005	6.0E-5
	0.0100	0.0095	0.0105	0.0100	0.0100	6.0E-5
	0.0150	0.0145	0.0155	0.0150	0.0150	6.0E-5
	0.0200	0.0195	0.0205	0.0200	0.0200	6.0E-5
	0.0300	0.0295	0.0305	0.0300	0.0300	6.0E-5
Counter Clockwise	(in)	(in)	(in)	(in)	(in)	(in)
	0.0005	0.0000	0.0010	0.0005	0.0005	6.0E-5
	0.0100	0.0095	0.0105	0.0100	0.0100	6.0E-5
	0.0150	0.0145	0.0155	0.0150	0.0150	6.0E-5
	0.0200	0.0195	0.0205	0.0200	0.0200	6.0E-5
	0.0300	0.0295	0.0305	0.0300	0.0300	6.0E-5
Repeat	(in)	(in)	(in)	(in)	(in)	(in)
	0.0000	-0.0005	0.0005	0.0000	0.0000	6.0E-5

COMMENTS

Example certificate, actual results and uncertainties will be reported at the time of calibration.
 Nominal and limits may vary based on actual make/model.

Fox Valley Metrology

3114 Medalist Drive

Oshkosh, WI 54902

(920) 426-5894 • Fax (920) 426-8120

<https://www.FoxValleyMetrology.com>

CERTIFICATE OF CALIBRATION



Certificate No. ACT-1272

-
- This certificate shall not be altered in any form or reproduced, except in full, without prior written approval from originating lab. These results relate only to the item(s) calibrated. Form Revision 10: 06/04/2024
 - Total expanded measurement uncertainties expressed are based on a confidence level of 95%; coverage factor of (k=2). The statement of compliance in this certificate was issued without taking the uncertainty of measurement into consideration. The customer shall assess the results and uncertainty when determining if the results meet their needs. (This is considered "shared responsibility.") Uncertainties expressed in nominal units.
 - The calibrations within the certificate/report are traceable through NIST or another National Metrology Institute to the International System of Units (SI). Calibration was completed in accordance with ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. Other standards listed upon request.
-