

**CERTIFICATE OF  
 CALIBRATION**



Certificate No. ACT-1272

<b>CERTIFICATION NUMBER</b> CP134-00334-840 <b>FOR</b> Sample Manufacturing 3121 Medalist Drive Oshkosh, WI 54902 <b>TEST INSTRUMENT</b> Recess Penetration Gage Dial Indicator/Hex Point <b>MAKE</b> Standard ASME B18.3 <b>MODEL</b> Dial/#9025-15 <b>RANGE</b> 5/32" <b>CUSTOMER LOCATION</b> Location1 Location2 <b>CONDITION RECEIVED</b> In Tolerance <b>CONDITION RETURNED</b> In Tolerance <b>CALIBRATED BY</b> Brittany Lemke <b>REVIEWED/ISSUED</b> D.P. 07/29/2025 <b>CALIBRATION LOCATION</b> FVM <b>ENVIRONMENT</b> 69.0°F, 31.0%RH, 28.92inHg <b>CALIBRATION DATE</b> 05/14/2025 <b>RECALIBRATION DUE</b> 05/14/2026	<b>IDENTIFICATION</b> HG-023B <b>SERIAL NUMBER</b> 334 <b>PURCHASE ORDER #</b> <b>PROCEDURES FOLLOWED</b> HG-023B rev. 1 HT-020B rev. 1 <b>STANDARDS USED</b> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>INSTRUMENT</th> <th>SERIAL NUMBER</th> <th>TRACE NUMBER</th> <th>NEXT CAL</th> </tr> </thead> <tbody> <tr> <td>STS-025</td> <td></td> <td>C0320-39582-705</td> <td>07/31/2025</td> </tr> <tr> <td>STS-135</td> <td>16782535.2</td> <td>C0172-20499-690</td> <td>07/31/2025</td> </tr> <tr> <td>STS-216</td> <td>81905.1</td> <td>CP064-27398-813</td> <td>07/31/2025</td> </tr> <tr> <td>STS-316A</td> <td></td> <td>C0221-46850-511</td> <td>07/31/2025</td> </tr> </tbody> </table>	INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL	STS-025		C0320-39582-705	07/31/2025	STS-135	16782535.2	C0172-20499-690	07/31/2025	STS-216	81905.1	CP064-27398-813	07/31/2025	STS-316A		C0221-46850-511	07/31/2025
INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL																		
STS-025		C0320-39582-705	07/31/2025																		
STS-135	16782535.2	C0172-20499-690	07/31/2025																		
STS-216	81905.1	CP064-27398-813	07/31/2025																		
STS-316A		C0221-46850-511	07/31/2025																		



**CALIBRATION RESULTS**

\* DENOTES "OUT OF TOLERANCE"

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
Hex Point	(in)	(in)	(in)	(in)	(in)	(in)
Across Flats						
Width (A)	0.1562	0.1562	0.1564	0.1562	0.1562	6.0E-5
Width (A)	0.1562	0.1562	0.1564	0.1562	0.1562	6.0E-5
Width (A)	0.1562	0.1562	0.1564	0.1562	0.1562	6.0E-5
Across Corners						
Width (B)	0.1774	0.1774	0.1776	0.1774	0.1774	6.0E-5
Width (B)	0.1774	0.1774	0.1776	0.1774	0.1774	6.0E-5
Width (B)	0.1774	0.1774	0.1776	0.1774	0.1774	6.0E-5

**COMMENTS**

Example certificate, actual results and uncertainties will be reported at the time of calibration. Nominal and limits may vary based on actual make/model.

- This certificate shall not be altered in any form or reproduced, except in full, without prior written approval from originating lab. These results relate only to the item(s) calibrated. Form Revision 10: 06/04/2024
- Total expanded measurement uncertainties expressed are based on a confidence level of 95%; coverage factor of (k=2). The statement of compliance in this certificate was issued without taking the uncertainty of measurement into consideration. The customer shall assess the results and uncertainty when determining if the results meet their needs. (This is considered "shared responsibility.") Uncertainties expressed in nominal units.
- The calibrations within the certificate/report are traceable through NIST or another National Metrology Institute to the International System of Units (SI). Calibration was completed in accordance with ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. Other standards listed upon request.