

**CERTIFICATE OF
 CALIBRATION**



Certificate No. ACT-1272

<p>CERTIFICATION NUMBER CP134-00190-840</p> <p>FOR Sample Manufacturing 3121 Medalist Drive Oshkosh, WI 54902</p> <p>TEST INSTRUMENT Scales Digital</p> <p>MAKE Standard Scale</p> <p>MODEL Standard Scale</p> <p>RANGE 60 lb</p> <p>CUSTOMER LOCATION Location1 Location2</p> <p>CONDITION RECEIVED In Tolerance</p> <p>CONDITION RETURNED In Tolerance</p> <p>CALIBRATED BY Alex Paulsen</p> <p>REVIEWED/ISSUED D.P. 07/29/2025</p> <p>CALIBRATION LOCATION FVM</p> <p>ENVIRONMENT 69.0°F, 31.0%RH, 28.92inHg</p> <p>CALIBRATION DATE 05/14/2025</p> <p>RECALIBRATION DUE 05/14/2026</p>	<p>IDENTIFICATION MC-009A</p> <p>SERIAL NUMBER 190</p> <p>PURCHASE ORDER #</p> <p>PROCEDURES FOLLOWED MC-009A rev. 3</p> <p>STANDARDS USED</p> <table border="1"> <thead> <tr> <th>INSTRUMENT</th> <th>SERIAL NUMBER</th> <th>TRACE NUMBER</th> <th>NEXT CAL</th> </tr> </thead> <tbody> <tr> <td>FVB-123</td> <td>SYS1D08113483</td> <td>C0157-40485-591</td> <td>06/30/2025</td> </tr> <tr> <td>FVS-019G-YELLOW</td> <td>N/A</td> <td>CP006-48963-489</td> <td>07/31/2025</td> </tr> </tbody> </table>	INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL	FVB-123	SYS1D08113483	C0157-40485-591	06/30/2025	FVS-019G-YELLOW	N/A	CP006-48963-489	07/31/2025
INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL										
FVB-123	SYS1D08113483	C0157-40485-591	06/30/2025										
FVS-019G-YELLOW	N/A	CP006-48963-489	07/31/2025										



CALIBRATION RESULTS

* DENOTES "OUT OF TOLERANCE"

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
Increasing Load	(1b)	(1b)	(1b)	(1b)	(1b)	(1b)
	0.000	-0.005	0.005	0.000	0.000	1.4E-3
	1.000	0.995	1.005	1.000	1.000	1.5E-3
	2.000	1.995	2.005	2.000	2.000	1.5E-3
	5.000	4.995	5.005	5.000	5.000	1.5E-3
	10.000	9.995	10.005	10.000	10.000	1.5E-3
	20.000	19.995	20.005	20.000	20.000	2.2E-3
Capacity (1b)	50.000	49.990	50.010	50.000	50.000	2.7E-3
	60.000	59.990	60.010	60.000	60.000	2.8E-3
Decreasing Load	(1b)	(1b)	(1b)	(1b)	(1b)	(1b)
	30.000	29.990	30.010	30.000	30.000	2.3E-3
Shift Test	(1b)	(1b)	(1b)	(1b)	(1b)	(1b)
Position 1	20.000	19.995	20.005	20.000	20.000	2.2E-3
Position 2	20.000	19.995	20.005	20.000	20.000	2.2E-3
Position 3	20.000	19.995	20.005	20.000	20.000	2.2E-3
Position 4	20.000	19.995	20.005	20.000	20.000	2.2E-3

COMMENTS

Scale Capacity = 60 lb; Precision = 0.005 lb; Class = II; Total Divisions = 12000
 Example certificate, actual results and uncertainties will be reported at the time of calibration.
 Nominal and limits may vary based on actual make/model.

Fox Valley Metrology

3114 Medalist Drive

Oshkosh, WI 54902

(920) 426-5894 • Fax (920) 426-8120

<https://www.FoxValleyMetrology.com>

CERTIFICATE OF CALIBRATION



-
- This certificate shall not be altered in any form or reproduced, except in full, without prior written approval from originating lab. These results relate only to the item(s) calibrated. Form Revision 10: 06/04/2024
 - Total expanded measurement uncertainties expressed are based on a confidence level of 95%; coverage factor of (k=2). The statement of compliance in this certificate was issued without taking the uncertainty of measurement into consideration. The customer shall assess the results and uncertainty when determining if the results meet their needs. (This is considered "shared responsibility.") Uncertainties expressed in nominal units.
 - The calibrations within the certificate/report are traceable through NIST or another National Metrology Institute to the International System of Units (SI). Calibration was completed in accordance with ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. Other standards listed upon request.
-