

**CERTIFICATE OF
 CALIBRATION**



Certificate No. ACT-1272

CERTIFICATION NUMBER CP134-00122-840 FOR Sample Manufacturing 3121 Medalist Drive Oshkosh, WI 54902 TEST INSTRUMENT Tachometer MAKE Extech Instruments MODEL 461891 RANGE (0.5 to 19999) rpm CUSTOMER LOCATION Location1 Location2 CONDITION RECEIVED In Tolerance CONDITION RETURNED In Tolerance CALIBRATED BY Codi Pijan REVIEWED/ISSUED D.P. 07/29/2025 CALIBRATION LOCATION FVM ENVIRONMENT 69.0°F, 31.0%RH, 28.92inHg CALIBRATION DATE 05/14/2025 RECALIBRATION DUE 05/14/2026	IDENTIFICATION EL-009A SERIAL NUMBER 122 PURCHASE ORDER # PROCEDURES FOLLOWED EL-009A rev. 1 STANDARDS USED <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>INSTRUMENT</th> <th>SERIAL NUMBER</th> <th>TRACE NUMBER</th> <th>NEXT CAL</th> </tr> </thead> <tbody> <tr> <td>FVS-648</td> <td></td> <td>CP037-43890-690</td> <td>02/28/2026</td> </tr> <tr> <td>FVS-A49</td> <td>26818</td> <td>C0352-39128-917</td> <td>03/31/2026</td> </tr> </tbody> </table>	INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL	FVS-648		CP037-43890-690	02/28/2026	FVS-A49	26818	C0352-39128-917	03/31/2026
INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL										
FVS-648		CP037-43890-690	02/28/2026										
FVS-A49	26818	C0352-39128-917	03/31/2026										



CALIBRATION RESULTS

* DENOTES "OUT OF TOLERANCE"

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
Revolutions per Minute	(rpm)	(rpm)	(rpm)	(rpm)	(rpm)	(rpm)
	100.0	99.8	100.2	100.1	100.1	8.0E-2
	500.0	499.6	500.4	500.1	500.1	4.0E-1
	1000	998	1002	1000	1000	8.0E-1
	2000	1998	2002	2000	2000	1.6E+0
	3000	2997	3003	3000	3000	2.4E+0
	4000	3997	4003	4000	4000	3.2E+0
	5000	4996	5004	5000	5000	4.0E+0
	6000	5996	6004	6000	6000	4.8E+0

COMMENTS

Example certificate, actual results and uncertainties will be reported at the time of calibration. Nominal and limits may vary based on actual make/model.

- This certificate shall not be altered in any form or reproduced, except in full, without prior written approval from originating lab. These results relate only to the item(s) calibrated. Form Revision 10: 06/04/2024
- Total expanded measurement uncertainties expressed are based on a confidence level of 95%; coverage factor of (k=2). The statement of compliance in this certificate was issued without taking the uncertainty of measurement into consideration. The customer shall assess the results and uncertainty when determining if the results meet their needs. (This is considered "shared responsibility.") Uncertainties expressed in nominal units.
- The calibrations within the certificate/report are traceable through NIST or another National Metrology Institute to the International System of Units (SI). Calibration was completed in accordance with ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. Other standards listed upon request.