

**CERTIFICATE OF
 CALIBRATION**



Certificate No. ACT-1272

CERTIFICATION NUMBER	CP134-00027-840	IDENTIFICATION	HG-015A
FOR	Sample Manufacturing 3121 Medalist Drive Oshkosh, WI 54902	SERIAL NUMBER	27
TEST INSTRUMENT	Telescope Gages	PURCHASE ORDER #	
MAKE	Mitutoyo	PROCEDURES FOLLOWED	HG-015A rev. 1
MODEL	155-903		
RANGE	2.00000 in	STANDARDS USED	
CUSTOMER LOCATION	Location1 Location2	INSTRUMENT	SERIAL NUMBER
CONDITION RECEIVED	In Tolerance	T-0022	271277
CONDITION RETURNED	In Tolerance	T-0540	601
CALIBRATED BY	Laura Fuhrmann	T-0697	TG2103-1071
REVIEWED/ISSUED	D.P. 07/29/2025	T-0919	16782940.77
CALIBRATION LOCATION	FVM	TRACE NUMBER	NEXT CAL
ENVIRONMENT	69.0°F, 31.0%RH, 28.92inHg	CP014-47951-847	01/31/2026
CALIBRATION DATE	05/14/2025	C0277-33085-514	09/30/2025
RECALIBRATION DUE	05/14/2026	CP030-55139-844	01/31/2026
		CP0151-17144-690	05/31/2025



CALIBRATION RESULTS

* DENOTES "OUT OF TOLERANCE"

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
Length	(in)	(in)	(in)	(in)	(in)	(in)
	0.50000			0.49010	0.49010	4.1E-5
	0.75000			0.77602	0.77602	4.2E-5
	(in)	(in)	(in)	(in)	(in)	(in)
	2.12500			2.12666	2.12666	4.4E-5
	3.50000			3.66501	3.66501	4.7E-5
	(in)	(in)	(in)	(in)	(in)	(in)
	0.75000			0.75435	0.75435	4.1E-5
	1.25000			1.29480	1.29480	4.3E-5
	(in)	(in)	(in)	(in)	(in)	(in)
	3.50000			3.46741	3.46741	4.7E-5
	6.00000			3.12945	3.12945	4.6E-5
	(in)	(in)	(in)	(in)	(in)	(in)
	0.31250			0.30062	0.30062	4.1E-5
	0.50000			0.51027	0.51027	4.1E-5
	(in)	(in)	(in)	(in)	(in)	(in)
	1.25000			1.26062	1.26062	4.2E-5
	2.12500			2.17730	2.17730	4.4E-5

COMMENTS

Fox Valley Metrology

3114 Medalist Drive

Oshkosh, WI 54902

(920) 426-5894 • Fax (920) 426-8120

<https://www.FoxValleyMetrology.com>

CERTIFICATE OF CALIBRATION



Certificate No. ACT-1272

Example certificate, actual results and uncertainties will be reported at the time of calibration.

Nominal and limits may vary based on actual make/model.

-
- This certificate shall not be altered in any form or reproduced, except in full, without prior written approval from originating lab. These results relate only to the item(s) calibrated. Form Revision 10: 06/04/2024
 - Total expanded measurement uncertainties expressed are based on a confidence level of 95%; coverage factor of (k=2). The statement of compliance in this certificate was issued without taking the uncertainty of measurement into consideration. The customer shall assess the results and uncertainty when determining if the results meet their needs. (This is considered "shared responsibility.") Uncertainties expressed in nominal units.
 - The calibrations within the certificate/report are traceable through NIST or another National Metrology Institute to the International System of Units (SI). Calibration was completed in accordance with ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. Other standards listed upon request.