Fox Valley Metrology
3114 Medalist Drive
Oshkosh, WI 54902
(920) 426-5894 • Fax (920) 426-8120
https://www.FoxValleyMetrology.com

CERTIFICATE OF CALIBRATION

ilac-MRA
ANAB
ANSI National Accreditation Board
ACCREDITED
ISO/IEC 17025
CALIBRATION AND DIMENSIONAL MEASUREMENT
Certificate No. ACT-1272

CERTIFICATION NUMBER	CP134-00297-840
FOR	Sample Manufacturing 3121 Medalist Drive Oshkosh, WI 54902
TEST INSTRUMENT	Transducer Current
MAKE	DaniSense
MODEL	DT100ID
RANGE	(0 to 100) ACI,
ACCURACY	±0.01% reading +.0.0002%FS
CUSTOMER LOCATION	Location1 Location2
CONDITION RECEIVED	In Tolerance
CONDITION RETURNED	In Tolerance
CALIBRATED BY	Brian L Gliszinski
REVIEWED/ISSUED	D.P. 07/29/2025
CALIBRATION LOCATION	FVM
ENVIRONMENT	69.0°F, 31.0%RH, 28.92inHg
CALIBRATION DATE	05/14/2025
RECALIBRATION DUE	05/14/2026

IDENTIFICATION	EL-007B
SERIAL NUMBER	297
PURCHASE ORDER #	

PROCEDURES FOLLOWED

EL-007B rev. 0

STANDARDS USED

INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL
FVS-275E	16783134	CO134-20963-690	05/31/2025
FVS-345	2823A24988	CO232-35013-823	08/31/2025
FVS-581	3025701	CP098-20905-524	04/30/2026
FVS-A40	6683603	CO309-40217-524	10/31/2025

CALIBRATION RESULTS

* DENOTES "OUT OF TOLERANCE"

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
Current	(mA)	(mA)	(mA)	(mA)	(mA)	(mA)
20A 1kHz	20.000	19.980	20.020	20.001	20.001	6.0E-4
40A 1kHz	40.000	39.940	40.060	40.000	40.000	6.0E-4
60A 1kHz	60.000	59.920	60.080	60.002	60.002	6.0E-4
80A 1kHz	80.000	79.900	80.100	80.001	80.001	6.0E-4
100A 1kHz	100.00	99.88	100.12	100.01	100.01	6.0E-3

Reviewed

COMMENTS

Example certificate, actual results and uncertainties will be reported at the time of calibration.
Nominal and limits may vary based on actual make/model.

■ This certificate shall not be altered in any form or reproduced, except in full, without prior written approval from originating lab. These results relate only to the item(s) calibrated. Form Revision 10: 06/04/2024
■ Total expanded measurement uncertainties expressed are based on a confidence level of 95%; coverage factor of (k=2). The statement of compliance in this certificate was issued without taking the uncertainty of measurement into consideration. The customer shall assess the results and uncertainty when determining if the results meet their needs. (This is considered “shared responsibility.”) Uncertainties expressed in nominal units.
■ The calibrations within the certificate/report are traceable through NIST or another National Metrology Institute to the International System of Units (SI). Calibration was completed in accordance with ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. Other standards listed upon request.