

**CERTIFICATE OF
 CALIBRATION**



Certificate No. ACT-1272

<p>CERTIFICATION NUMBER CP134-00197-840</p> <p>FOR Sample Manufacturing 3121 Medalist Drive Oshkosh, WI 54902</p> <p>TEST INSTRUMENT Weights</p> <p>MAKE Standard Weight MODEL Standard Weight</p> <p>CUSTOMER LOCATION Location1 Location2</p> <p>CONDITION RECEIVED In Tolerance CONDITION RETURNED In Tolerance</p> <p>CALIBRATED BY Garrett Beachum REVIEWED/ISSUED D.P. 07/29/2025 CALIBRATION LOCATION FVM ENVIRONMENT 69.0°F, 31.0%RH, 28.92inHg CALIBRATION DATE 05/14/2025 RECALIBRATION DUE 05/14/2026</p>	<p>IDENTIFICATION MC-012A</p> <p>SERIAL NUMBER 197</p> <p>PURCHASE ORDER #</p> <p>PROCEDURES FOLLOWED MC-012A rev. 2</p> <p>STANDARDS USED</p> <table border="1"> <thead> <tr> <th>INSTRUMENT</th> <th>SERIAL NUMBER</th> <th>TRACE NUMBER</th> <th>NEXT CAL</th> </tr> </thead> <tbody> <tr> <td>FVS-019V</td> <td></td> <td>C0290-28623-524</td> <td>10/31/2026</td> </tr> <tr> <td>FVS-098</td> <td>F1161200520519</td> <td>CP121-51996-681</td> <td>11/30/2025</td> </tr> <tr> <td>FVS-114</td> <td>13209856</td> <td>CP007-25831-379</td> <td>07/31/2025</td> </tr> <tr> <td>FVS-275M</td> <td>16787403.28</td> <td>C0141-16400-690</td> <td>05/31/2025</td> </tr> </tbody> </table>	INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL	FVS-019V		C0290-28623-524	10/31/2026	FVS-098	F1161200520519	CP121-51996-681	11/30/2025	FVS-114	13209856	CP007-25831-379	07/31/2025	FVS-275M	16787403.28	C0141-16400-690	05/31/2025
INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL																		
FVS-019V		C0290-28623-524	10/31/2026																		
FVS-098	F1161200520519	CP121-51996-681	11/30/2025																		
FVS-114	13209856	CP007-25831-379	07/31/2025																		
FVS-275M	16787403.28	C0141-16400-690	05/31/2025																		



CALIBRATION RESULTS

* DENOTES "OUT OF TOLERANCE"

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
NIST 105-1 Class F Conventional Mass				FVS-114	FVS-114	
Correction	(g)	(g)	(g)	(g)	(g)	(g)
1 kg	0.00	-0.10	0.10	0.00	0.00	6.0E-3

COMMENTS

Test Weights Conventional Mass Correction Reference = C0003-WGHTS-FVM
 Example certificate, actual results and uncertainties will be reported at the time of calibration.
 Nominal and limits may vary based on actual make/model.

- This certificate shall not be altered in any form or reproduced, except in full, without prior written approval from originating lab. These results relate only to the item(s) calibrated. Form Revision 10: 06/04/2024
- Total expanded measurement uncertainties expressed are based on a confidence level of 95%; coverage factor of (k=2). The statement of compliance in this certificate was issued without taking the uncertainty of measurement into consideration. The customer shall assess the results and uncertainty when determining if the results meet their needs. (This is considered "shared responsibility.") Uncertainties expressed in nominal units.
- The calibrations within the certificate/report are traceable through NIST or another National Metrology Institute to the International System of Units (SI). Calibration was completed in accordance with ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. Other standards listed upon request.